

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111824\
 Data File : VD080071.D
 Acq On : 18 Nov 2024 16:12
 Operator : RP/MD
 Sample : P4850-01
 Misc : 5.04G/5ml/MSVOA_D/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-03-11142024

Quant Time: Nov 19 04:48:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 15 16:25:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	2953	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.781	114	4817	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	4484	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	1533	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.222	65	2083	68.511	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 137.020%	
35) Dibromofluoromethane	7.810	113	2228	69.497	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 139.000%	
50) Toluene-d8	10.269	98	4746	39.279	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 78.560%	
62) 4-Bromofluorobenzene	12.575	95	1599	40.335	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 80.680%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111824\
Data File : VD080071.D
Acq On : 18 Nov 2024 16:12
Operator : RP/MD
Sample : P4850-01
Misc : 5.04G/5ml/MSVOA_D/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-03-11142024

Quant Time: Nov 19 04:48:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111524S.M
Quant Title : SW846 8260
QLast Update : Fri Nov 15 16:25:47 2024
Response via : Initial Calibration

