

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111824\
 Data File : VD080076.D
 Acq On : 18 Nov 2024 18:28
 Operator : RP/MD
 Sample : P4860-05
 Misc : 6.25G/5ml/MSVOA_D/SOIL/A
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PH2-BOT-004

Quant Time: Nov 19 04:50:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111524S.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 15 16:25:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	200921	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.780	114	364222	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.580	117	324341	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	143861	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.233	65	84226	40.715	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	81.420%
35) Dibromofluoromethane	7.810	113	91562	37.773	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	75.540%
50) Toluene-d8	10.269	98	290784	31.828	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	63.660%
62) 4-Bromofluorobenzene	12.574	95	91664	30.581	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	61.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111824\
Data File : VD080076.D
Acq On : 18 Nov 2024 18:28
Operator : RP/MD
Sample : P4860-05
Misc : 6.25G/5ml/MSVOA_D/SOIL/A
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PH2-BOT-004

Quant Time: Nov 19 04:50:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111524S.M
Quant Title : SW846 8260
QLast Update : Fri Nov 15 16:25:47 2024
Response via : Initial Calibration

