

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD112020\
 Data File : VD067734.D
 Acq On : 20 Nov 2020 19:32
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 21 01:36:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D110320S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 00:56:28 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	72	0.00
2 T	Dichlorodifluoromethane	0.457	0.366	19.9	64	0.00
3 P	Chloromethane	0.311	0.226	27.3#	59	0.00
4 C	Vinyl Chloride	0.347	0.272	21.6#	59	0.00
5 T	Bromomethane	0.281	0.234	16.7	64	0.00
6 T	Chloroethane	0.214	0.186	13.1	63	-0.01
7 T	Trichlorofluoromethane	0.899	0.837	6.9	71	0.00
8 T	Diethyl Ether	0.181	0.190	-5.0	79	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.454	0.399	12.1	66	0.00
10 T	Methyl Iodide	0.486	0.413	15.0	57	0.00
11 T	Tert butyl alcohol	0.022	0.025	-13.6	97	0.00
12 CM	1,1-Dichloroethene	0.417	0.378	9.4#	69	0.00
13 T	Acrolein	0.016	0.017	-6.3	63	-0.01
14 T	Allyl chloride	0.447	0.393	12.1	64	0.00
15 T	Acrylonitrile	0.067	0.071	-6.0	77	-0.01
16 T	Acetone	0.065	0.059	9.2	58	0.00
17 T	Carbon Disulfide	1.279	1.024	19.9	60	0.00
18 T	Methyl Acetate	0.135	0.152	-12.6	91	0.00
19 T	Methyl tert-butyl Ether	0.823	0.951	-15.6	83	0.00
20 T	Methylene Chloride	0.462	0.468	-1.3	79	0.00
21 T	trans-1,2-Dichloroethene	0.486	0.466	4.1	71	0.00
22 T	Diisopropyl ether	0.870	0.869	0.1	72	0.00
23 T	Vinyl Acetate	0.456	0.485	-6.4	73	0.00
24 P	1,1-Dichloroethane	0.712	0.709	0.4	75	-0.01
25 T	2-Butanone	0.070	0.077	-10.0	80	0.00
26 T	2,2-Dichloropropane	0.751	0.730	2.8	72	0.00
27 T	cis-1,2-Dichloroethene	0.505	0.518	-2.6	74	0.00
28 T	Bromochloromethane	0.253	0.255	-0.8	72	0.00
29 T	Tetrahydrofuran	0.042	0.046	-9.5	78	-0.01
30 C	Chloroform	0.849	0.903	-6.4#	78	0.00
31 T	Cyclohexane	0.634	0.504	20.5	61	-0.01
32 T	1,1,1-Trichloroethane	0.876	0.896	-2.3	77	0.00
33 S	1,2-Dichloroethane-d4	0.437	0.433	0.9	80	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	74	0.00
35 S	Dibromofluoromethane	0.301	0.274	9.0	71	0.00
36 T	1,1-Dichloropropene	0.428	0.398	7.0	69	-0.01
37 T	Ethyl Acetate	0.111	0.113	-1.8	71	0.00
38 T	Carbon Tetrachloride	0.556	0.566	-1.8	76	0.00
39 T	Methylcyclohexane	0.504	0.455	9.7	63	0.00
40 TM	Benzene	1.159	1.147	1.0	73	0.00
41 T	Methacrylonitrile	0.066	0.071	-7.6	86	-0.01
42 TM	1,2-Dichloroethane	0.358	0.400	-11.7	84	0.00
43 T	Isopropyl Acetate	0.220	0.235	-6.8	79	0.00
44 TM	Trichloroethene	0.381	0.380	0.3	76	0.00
45 C	1,2-Dichloropropane	0.251	0.244	2.8#	73	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD112020\
 Data File : VD067734.D
 Acq On : 20 Nov 2020 19:32
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 21 01:36:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D110320S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 00:56:28 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.159	0.173	-8.8	80	0.00
47 T	Bromodichloromethane	0.433	0.472	-9.0	80	0.00
48 T	Methyl methacrylate	0.113	0.115	-1.8	72	0.00
49 T	1,4-Dioxane	0.002	0.002	0.0	86	0.00
50 S	Toluene-d8	1.059	0.918	13.3	68	0.00
51 T	4-Methyl-2-Pentanone	0.105	0.118	-12.4	80	0.00
52 CM	Toluene	0.777	0.807	-3.9#	76	0.00
53 T	t-1,3-Dichloropropene	0.371	0.409	-10.2	81	0.00
54 T	cis-1,3-Dichloropropene	0.441	0.457	-3.6	77	0.00
55 T	1,1,2-Trichloroethane	0.211	0.232	-10.0	81	0.00
56 T	Ethyl methacrylate	0.218	0.246	-12.8	81	0.00
57 T	1,3-Dichloropropane	0.339	0.358	-5.6	77	0.00
58 T	2-Chloroethyl Vinyl ether	0.102	0.113	-10.8	77	0.00
59 T	2-Hexanone	0.073	0.079	-8.2	74	0.00
60 T	Dibromochloromethane	0.312	0.353	-13.1	83	0.00
61 T	1,2-Dibromoethane	0.212	0.223	-5.2	79	0.00
62 S	4-Bromofluorobenzene	0.365	0.351	3.8	75	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	75	0.00
64 T	Tetrachloroethene	0.370	0.363	1.9	74	0.00
65 PM	Chlorobenzene	0.963	0.988	-2.6	78	0.00
66 T	1,1,1,2-Tetrachloroethane	0.380	0.409	-7.6	82	0.00
67 C	Ethyl Benzene	1.648	1.686	-2.3#	75	0.00
68 T	m/p-Xylenes	0.641	0.656	-2.3	75	0.00
69 T	o-Xylene	0.572	0.610	-6.6	78	0.00
70 T	Styrene	0.998	1.052	-5.4	77	0.00
71 P	Bromoform	0.194	0.223	-14.9	87	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	78	0.00
73 T	Isopropylbenzene	3.231	3.250	-0.6	74	0.00
74 T	N-amyl acetate	0.425	0.428	-0.7	76	0.00
75 P	1,1,2,2-Tetrachloroethane	0.441	0.449	-1.8	80	0.00
76 T	1,2,3-Trichloropropane	0.302	0.274	9.3	76	0.00
77 T	Bromobenzene	0.794	0.826	-4.0	80	0.00
78 T	n-propylbenzene	3.696	3.603	2.5	72	0.00
79 T	2-Chlorotoluene	2.100	2.139	-1.9	77	0.00
80 T	1,3,5-Trimethylbenzene	2.760	2.846	-3.1	76	0.00
81 T	trans-1,4-Dichloro-2-butene	0.149	0.146	2.0	80	0.00
82 T	4-Chlorotoluene	2.252	2.268	-0.7	77	0.00
83 T	tert-Butylbenzene	2.339	2.358	-0.8	74	0.00
84 T	1,2,4-Trimethylbenzene	2.748	2.858	-4.0	77	0.00
85 T	sec-Butylbenzene	3.158	3.130	0.9	72	0.00
86 T	p-Isopropyltoluene	3.050	3.047	0.1	74	0.00
87 T	1,3-Dichlorobenzene	1.514	1.527	-0.9	78	0.00
88 T	1,4-Dichlorobenzene	1.533	1.534	-0.1	79	0.00
89 T	n-Butylbenzene	2.634	2.543	3.5	70	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD112020\
Data File : VD067734.D
Acq On : 20 Nov 2020 19:32
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Nov 21 01:36:59 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D110320S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 00:56:28 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.574	0.588	-2.4	78	0.00
91 T	1,2-Dichlorobenzene	1.311	1.358	-3.6	80	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.078	0.093	-19.2	95	0.00
93 T	1,2,4-Trichlorobenzene	0.890	0.946	-6.3	80	0.00
94 T	Hexachlorobutadiene	0.595	0.583	2.0	76	0.00
95 T	Naphthalene	1.373	1.640	-19.4	85	0.00
96 T	1,2,3-Trichlorobenzene	0.762	0.853	-11.9	83	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6