

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD112020\
 Data File : VD067733.D
 Acq On : 20 Nov 2020 19:05
 Operator : VA/SY
 Sample : L4847-03
 Misc : 1.09G/5.00ml/MSVOA D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EDISON-4A

Quant Time: Nov 21 01:34:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D110320S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 00:56:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	244197	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	364809	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	370160	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	181935	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	94369	44.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.40%	
35) Dibromofluoromethane	7.92	113	105175	47.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.70%	
50) Toluene-d8	10.34	98	428096	55.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.84%	
62) 4-Bromofluorobenzene	12.64	95	152799	57.30	ug/l	0.00
Spiked Amount	50.000		Recovery	=	114.60%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD112020\
Data File : VD067733.D
Acq On : 20 Nov 2020 19:05
Operator : VA/SY
Sample : L4847-03
Misc : 1.09G/5.00ml/MSVOA D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EDISON-4A

Quant Time: Nov 21 01:34:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D110320S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 00:56:28 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.98 min Scan# 1030

Delta R.T. -0.01 min

Lab File: VD067733.D

Acq: 20 Nov 2020 19:05

Instrument :

MSVOA_D

ClientSampleId :


EDISON-4A

Tot Ion:168 Resp: 244197

Ion Ratio Lower Upper




168 100

99 50.9 45.0 67.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VDC



#33

1,2-Dichloroethane-d4

Concen: 44.196 ug/l

RT: 8.33 min Scan# 1090

Delta R.T. -0.01 min

Lab File: VD067733.D


Acq: 20 Nov 2020 19:05

Tgt Ion: 65 Resp: 94369

Ion Ratio Lower Upper



65 100


67 47.3 0.0 99.4



Abundance Ion 64.90 (64.60 to 65.60): VDC

Ion 66.90 (66.60 to 67.60): VDC





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.87 min Scan# 1181

Delta R.T. -0.01 min

Lab File: VD067733.D

Acq: 20 Nov 2020 19:05

Instrument :

MSVOA_D

ClientSampleId :

EDISON-4A


Tot Ion:114 Resp: 364809

Ion Ratio Lower Upper

114 100

63 16.3 0.0 35.0




88 15.1 0.0 33.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): V

Ion 87.90 (87.60 to 88.60): V



#35

Dibromofluoromethane

Concen: 47.851 ug/l

RT: 7.92 min Scan# 1020

Delta R.T. -0.00 min

Lab File: VD067733.D

Acq: 20 Nov 2020 19:05


Tgt Ion:113 Resp: 105175

Ion Ratio Lower Upper

113 100

111 103.7 80.0 120.0



192 21.3 16.3 24.5




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 55.423 ug/l
RT: 10.34 min Scan# 1432
Delta R.T. -0.00 min
Lab File: VD067733.D
Acq: 20 Nov 2020 19:05

Instrument :
MSVOA_D
ClientSampleId :
EDISON-4A

Tot Ion: 98 Resp: 428096
Ion Ratio Lower Upper
98 100
100 64.7 53.0 79.6




Abundance Ion 98.00 (97.70 to 98.70): VDC
Ion 100.00 (99.70 to 100.70): VDC






#62
4-Bromofluorobenzene
Concen: 57.298 ug/l
RT: 12.64 min Scan# 1822
Delta R.T. -0.00 min
Lab File: VD067733.D
Acq: 20 Nov 2020 19:05

Tgt Ion: 95 Resp: 152799
Ion Ratio Lower Upper
95 100
174 97.3 0.0 177.8
176 92.9 0.0 172.0



Abundance Ion 94.90 (94.60 to 95.60): VDC
Ion 173.80 (173.50 to 174.50): VDC
Ion 175.80 (175.50 to 176.50): VDC









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.65 min Scan# 1654
Delta R.T. -0.00 min
Lab File: VD067733.D
Acq: 20 Nov 2020 19:05

Instrument :
MSVOA_D
ClientSampleId :
EDISON-4A

Tot Ion	Ratio	Lower	Upper
117	100		
82	49.4	41.6	62.4
119	30.3	25.1	37.7



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.58 min Scan# 1982
Delta R.T. -0.01 min
Lab File: VD067733.D
Acq: 20 Nov 2020 19:05

Tgt Ion	Ratio	Lower	Upper
152	100		
115	61.0	29.3	88.0
150	159.6	0.0	338.2

