

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_D\DATA\VD112118\
 Data File : VD060445.D
 Acq On : 21 Nov 2018 16:29
 Operator : JC/SY
 Sample : J5768-05
 Misc : 8.23g/5ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OK-03-112018

Quant Time: Nov 22 03:34:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D110618S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 19 16:31:59 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.40	168	8350	50.00	ug/l	-0.02
34) 1,4-Difluorobenzene	7.45	114	16567	50.00	ug/l	-0.01
63) Chlorobenzene-d5	11.28	117	9124	50.00	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	13.37	152	5048	50.00	ug/l	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.78	65	7695	128.71	ug/l	-0.02
Spiked Amount				50.000		
			Recovery	=	257.42%	
35) Dibromofluoromethane	6.32	113	6843	55.39	ug/l	-0.01
Spiked Amount				50.000		
			Recovery	=	110.78%	
50) Toluene-d8	9.45	98	4657	12.88	ug/l	-0.01
Spiked Amount				50.000		
			Recovery	=	25.76%	
62) 4-Bromofluorobenzene	12.43	95	4830	36.87	ug/l	-0.01
Spiked Amount				50.000		
			Recovery	=	73.74%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
5) Bromomethane	2.18	94	503	32.82	ug/l #	2
10) Methyl Iodide	3.27	142	182	2.80	ug/l #	35
16) Acetone	3.16	43	14240	1373.09	ug/l	97
18) Methyl Acetate	3.55	43	193	8.73	ug/l #	63
20) Methylene Chloride	3.72	84	185	2.01	ug/l #	1
81) trans-1,4-Dichloro-2-buten	12.43	75	1798	104.59	ug/l #	15
95) Naphthalene	14.96	128	4767	34.49	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_D\DATA\VD112118\
Data File : VD060445.D
Acq On : 21 Nov 2018 16:29
Operator : JC/SY
Sample : J5768-05
Misc : 8.23g/5ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OK-03-112018

Quant Time: Nov 22 03:34:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D110618S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 19 16:31:59 2018
Response via : Initial Calibration

