

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD112320\
 Data File : VD067744.D
 Acq On : 23 Nov 2020 14:32
 Operator : VA/SY
 Sample : VD1123SBL01
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1123SBL01

Quant Time: Nov 24 01:03:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D112320S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 24 00:51:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	277529	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	423925	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	380105	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	186414	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.33	65	144205	53.44	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.88%
35) Dibromofluoromethane	7.92	113	132702	49.19	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.38%
50) Toluene-d8	10.34	98	482813	48.95	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.90%
62) 4-Bromofluorobenzene	12.64	95	161740	47.97	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.94%

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD112320\
Data File : VD067744.D
Acq On : 23 Nov 2020 14:32
Operator : VA/SY
Sample : VD1123SBL01
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1123SBL01

Quant Time: Nov 24 01:03:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D112320S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 24 00:51:35 2020
Response via : Initial Calibration

