

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD112320\
 Data File : VD067750.D
 Acq On : 23 Nov 2020 17:16
 Operator : VA/SY
 Sample : L4838-01
 Misc : 4.77G/5.00ml/MSVOA D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 KENS-B2-111920-0-2

Quant Time: Nov 24 01:13:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D112320S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 24 00:51:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	249018	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	372231	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	403667	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	213109	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.33	65	108078	44.64	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	89.28%
35) Dibromofluoromethane	7.91	113	112344	47.43	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.86%
50) Toluene-d8	10.34	98	435807	50.32	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.64%
62) 4-Bromofluorobenzene	12.64	95	180626	61.02	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	122.04%

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD112320\
Data File : VD067750.D
Acq On : 23 Nov 2020 17:16
Operator : VA/SY
Sample : L4838-01
Misc : 4.77G/5.00ml/MSVOA D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
KENS-B2-111920-0-2

Quant Time: Nov 24 01:13:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D112320S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 24 00:51:35 2020
Response via : Initial Calibration

