

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD112420\
 Data File : VD067766.D
 Acq On : 24 Nov 2020 13:52
 Operator : VA/SY
 Sample : L4840-02
 Misc : 4.26G/5.00ml/MSVOA D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 30740

Quant Time: Nov 25 02:55:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D112320S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 24 00:51:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	238738	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	372324	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	422433	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	241445	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.34	65	113908	49.07	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.14%
35) Dibromofluoromethane	7.92	113	117579	49.62	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.24%
50) Toluene-d8	10.34	98	444044	51.26	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.52%
62) 4-Bromofluorobenzene	12.64	95	200463	67.70	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	135.40%

Target Compounds					Qvalue
40) Benzene	8.36	78	18946	2.159 ug/l	94
52) Toluene	10.40	92	16441	2.762 ug/l	95
68) m/p-Xylenes	11.86	106	7129	1.278 ug/l	99
95) Naphthalene	15.40	128	100581	13.239 ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD112420\
Data File : VD067766.D
Acq On : 24 Nov 2020 13:52
Operator : VA/SY
Sample : L4840-02
Misc : 4.26G/5.00ml/MSVOA D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
30740

Quant Time: Nov 25 02:55:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D112320S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 24 00:51:35 2020
Response via : Initial Calibration

