

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD112420\
 Data File : VD067773.D
 Acq On : 24 Nov 2020 17:37
 Operator : VA/SY
 Sample : L4858-02
 Misc : 5.74G/5.00ml/MSVOA D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PHC-340EM-013

Quant Time: Nov 25 03:06:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D112320S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 24 00:51:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	45420	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	70489	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	69736	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	31322	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		8.33	65	40528	91.77	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	183.54%#	
35) Dibromofluoromethane		7.92	113	28833	64.27	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	128.54%	
50) Toluene-d8		10.34	98	75568	46.08	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	92.16%	
62) 4-Bromofluorobenzene		12.64	95	27039	48.23	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	96.46%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	4.16	43	8542	149.722	ug/l #	78
39) Methylcyclohexane	9.36	83	14779	20.646	ug/l #	89
80) 1,3,5-Trimethylbenzene	12.97	105	3036	1.694	ug/l	76
84) 1,2,4-Trimethylbenzene	13.27	105	7475	4.175	ug/l	96
95) Naphthalene	15.40	128	5980	6.068	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD112420\
Data File : VD067773.D
Acq On : 24 Nov 2020 17:37
Operator : VA/SY
Sample : L4858-02
Misc : 5.74G/5.00ml/MSVOA D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PHC-340EM-013

Quant Time: Nov 25 03:06:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D112320S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 24 00:51:35 2020
Response via : Initial Calibration

