

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD112520\
 Data File : VD067795.D
 Acq On : 25 Nov 2020 18:54
 Operator : VA/SY
 Sample : L4857-01RE
 Misc : 4.10G/5.00ml/MSVOA D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 304EM-STOCKPILERE

Quant Time: Nov 26 02:53:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D112320S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 24 00:51:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	160495	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	254073	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	205171	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	71041	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.33	65	98336	63.02	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	126.04%
35) Dibromofluoromethane	7.92	113	87243	53.96	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.92%
50) Toluene-d8	10.34	98	269374	45.57	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	91.14%
62) 4-Bromofluorobenzene	12.64	95	67547	33.43	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	66.86%

Target Compounds

					Qvalue
64) Tetrachloroethene	10.88	164	13188	8.282	ug/l 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD112520\
Data File : VD067795.D
Acq On : 25 Nov 2020 18:54
Operator : VA/SY
Sample : L4857-01RE
Misc : 4.10G/5.00ml/MSVOA D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
304EM-STOCKPILERE

Quant Time: Nov 26 02:53:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D112320S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 24 00:51:35 2020
Response via : Initial Calibration

