

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD112520\
 Data File : VD067798.D
 Acq On : 25 Nov 2020 20:14
 Operator : VA/SY
 Sample : L4858-01RE
 Misc : 2.95G/5.00ml/MSVOA D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PHC-340EM-012RE

Quant Time: Nov 26 03:01:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D112320S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 24 00:51:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.99	168	5875	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	10044	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	9209	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	4571	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.34	65	5803	101.59	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	203.18%#
35) Dibromofluoromethane	7.92	113	3544	55.44	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.88%
50) Toluene-d8	10.34	98	11114	47.56	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.12%
62) 4-Bromofluorobenzene	12.64	95	3992	49.98	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	99.96%

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD112520\
Data File : VD067798.D
Acq On : 25 Nov 2020 20:14
Operator : VA/SY
Sample : L4858-01RE
Misc : 2.95G/5.00ml/MSVOA D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PHC-340EM-012RE

Quant Time: Nov 26 03:01:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D112320S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 24 00:51:35 2020
Response via : Initial Calibration

