

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD112520\
 Data File : VD067802.D
 Acq On : 25 Nov 2020 22:03
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 26 03:08:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D112320S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 24 00:51:35 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	84	0.00
2 T	Dichlorodifluoromethane	0.435	0.437	-0.5	82	0.00
3 P	Chloromethane	0.242	0.254	-5.0	89	0.00
4 C	Vinyl Chloride	0.304	0.301	1.0#	85	0.00
5 T	Bromomethane	0.259	0.282	-8.9	97	0.00
6 T	Chloroethane	0.202	0.215	-6.4	93	0.01
7 T	Trichlorofluoromethane	0.926	0.916	1.1	84	0.00
8 T	Diethyl Ether	0.187	0.194	-3.7	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.444	0.420	5.4	83	0.00
10 T	Methyl Iodide	0.454	0.487	-7.3	84	0.00
11 T	Tert butyl alcohol	0.035	0.028	20.0	97	0.00
12 CM	1,1-Dichloroethene	0.410	0.410	0.0	87	0.00
13 T	Acrolein	0.020	0.017	15.0	78	0.00
14 T	Allyl chloride	0.397	0.422	-6.3	89	0.00
15 T	Acrylonitrile	0.065	0.070	-7.7	93	0.00
16 T	Acetone	0.063	0.060	4.8	82	0.00
17 T	Carbon Disulfide	1.187	1.163	2.0	83	0.00
18 T	Methyl Acetate	0.126	0.140	-11.1	99	0.00
19 T	Methyl tert-butyl Ether	0.877	0.988	-12.7	96	0.00
20 T	Methylene Chloride	0.630	0.524	16.8	94	0.00
21 T	trans-1,2-Dichloroethene	0.467	0.485	-3.9	90	0.00
22 T	Diisopropyl ether	0.783	0.854	-9.1	89	0.00
23 T	Vinyl Acetate	0.442	0.494	-11.8	91	0.00
24 P	1,1-Dichloroethane	0.664	0.715	-7.7	91	0.00
25 T	2-Butanone	0.070	0.072	-2.9	91	0.00
26 T	2,2-Dichloropropane	0.783	0.766	2.2	85	0.00
27 T	cis-1,2-Dichloroethene	0.497	0.544	-9.5	94	0.00
28 T	Bromochloromethane	0.213	0.237	-11.3	93	0.00
29 T	Tetrahydrofuran	0.040	0.044	-10.0	91	0.00
30 C	Chloroform	0.868	0.937	-7.9#	95	0.00
31 T	Cyclohexane	0.588	0.520	11.6	79	0.00
32 T	1,1,1-Trichloroethane	0.920	0.942	-2.4	89	0.00
33 S	1,2-Dichloroethane-d4	0.486	0.543	-11.7	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.318	0.349	-9.7	96	0.00
36 T	1,1-Dichloropropene	0.429	0.417	2.8	88	0.00
37 T	Ethyl Acetate	0.115	0.115	0.0	89	-0.01
38 T	Carbon Tetrachloride	0.622	0.603	3.1	87	0.00
39 T	Methylcyclohexane	0.508	0.487	4.1	83	0.00
40 TM	Benzene	1.178	1.183	-0.4	91	0.00
41 T	Methacrylonitrile	0.059	0.061	-3.4	88	0.00
42 TM	1,2-Dichloroethane	0.400	0.411	-2.7	95	0.00
43 T	Isopropyl Acetate	0.230	0.233	-1.3	92	0.00
44 TM	Trichloroethene	0.391	0.395	-1.0	92	0.00
45 C	1,2-Dichloropropane	0.242	0.245	-1.2#	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD112520\
 Data File : VD067802.D
 Acq On : 25 Nov 2020 22:03
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 26 03:08:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D112320S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 24 00:51:35 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.165	0.169	-2.4	95	0.00
47 T	Bromodichloromethane	0.458	0.483	-5.5	95	0.00
48 T	Methyl methacrylate	0.115	0.110	4.3	89	0.00
49 T	1,4-Dioxane	0.002	0.002	0.0	95	0.00
50 S	Toluene-d8	1.163	1.229	-5.7	93	0.00
51 T	4-Methyl-2-Pentanone	0.108	0.110	-1.9	92	0.00
52 CM	Toluene	0.799	0.830	-3.9#	93	0.00
53 T	t-1,3-Dichloropropene	0.400	0.414	-3.5	93	0.00
54 T	cis-1,3-Dichloropropene	0.453	0.455	-0.4	93	0.00
55 T	1,1,2-Trichloroethane	0.221	0.230	-4.1	96	0.00
56 T	Ethyl methacrylate	0.229	0.243	-6.1	95	0.00
57 T	1,3-Dichloropropane	0.348	0.368	-5.7	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.107	0.125	-16.8	96	0.00
59 T	2-Hexanone	0.077	0.077	0.0	91	0.00
60 T	Dibromochloromethane	0.337	0.350	-3.9	96	0.00
61 T	1,2-Dibromoethane	0.218	0.223	-2.3	93	0.00
62 S	4-Bromofluorobenzene	0.398	0.425	-6.8	95	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
64 T	Tetrachloroethene	0.388	0.374	3.6	92	0.00
65 PM	Chlorobenzene	0.983	1.000	-1.7	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.398	0.411	-3.3	97	0.00
67 C	Ethyl Benzene	1.697	1.698	-0.1#	91	0.00
68 T	m/p-Xylenes	0.660	0.666	-0.9	93	0.00
69 T	o-Xylene	0.601	0.604	-0.5	91	0.00
70 T	Styrene	1.024	1.069	-4.4	95	0.00
71 P	Bromoform	0.216	0.219	-1.4	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
73 T	Isopropylbenzene	3.290	3.413	-3.7	92	0.00
74 T	N-amyl acetate	0.434	0.440	-1.4	93	0.00
75 P	1,1,2,2-Tetrachloroethane	0.442	0.449	-1.6	98	0.00
76 T	1,2,3-Trichloropropane	0.348	0.363	-4.3	98	0.00
77 T	Bromobenzene	0.824	0.869	-5.5	97	0.00
78 T	n-propylbenzene	3.660	3.805	-4.0	92	0.00
79 T	2-Chlorotoluene	2.167	2.223	-2.6	92	0.00
80 T	1,3,5-Trimethylbenzene	2.861	3.005	-5.0	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.149	0.146	2.0	89	0.00
82 T	4-Chlorotoluene	2.309	2.370	-2.6	93	0.00
83 T	tert-Butylbenzene	2.459	2.561	-4.1	91	0.00
84 T	1,2,4-Trimethylbenzene	2.858	3.004	-5.1	93	0.00
85 T	sec-Butylbenzene	3.230	3.236	-0.2	89	0.00
86 T	p-Isopropyltoluene	3.129	3.161	-1.0	90	0.00
87 T	1,3-Dichlorobenzene	1.594	1.595	-0.1	95	0.00
88 T	1,4-Dichlorobenzene	1.579	1.547	2.0	90	0.00
89 T	n-Butylbenzene	2.681	2.658	0.9	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD112520\
Data File : VD067802.D
Acq On : 25 Nov 2020 22:03
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Nov 26 03:08:33 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D112320S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 24 00:51:35 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.596	0.587	1.5	90	0.00
91 T	1,2-Dichlorobenzene	1.373	1.392	-1.4	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.095	0.091	4.2	90	0.00
93 T	1,2,4-Trichlorobenzene	0.972	0.991	-2.0	91	0.00
94 T	Hexachlorobutadiene	0.621	0.631	-1.6	92	0.00
95 T	Naphthalene	1.573	1.623	-3.2	92	0.00
96 T	1,2,3-Trichlorobenzene	0.828	0.851	-2.8	91	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 5