

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD112619\
 Data File : VD064384.D
 Acq On : 26 Nov 2019 15:16
 Operator : VA/SY
 Sample : K6024-01
 Misc : 6.17G/5.00ml/MSVOA D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 B-1-(0.5-1.0)

Quant Time: Nov 27 06:18:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D110819S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 12 02:43:24 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.99	168	754863	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	1120076	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	972646	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	443070	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.34	65	313521	50.50	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.00%	
35) Dibromofluoromethane	7.92	113	334873	52.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.16%	
50) Toluene-d8	10.34	98	1302784	53.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.42%	
62) 4-Bromofluorobenzene	12.64	95	403239	48.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.02%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD112619\
Data File : VD064384.D
Acq On : 26 Nov 2019 15:16
Operator : VA/SY
Sample : K6024-01
Misc : 6.17G/5.00ml/MSVOA D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
B-1-(0.5-1.0)

Quant Time: Nov 27 06:18:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D110819S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 12 02:43:24 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.99 min Scan# 1031

Delta R.T. 0.00 min

Lab File: VD064384.D

Acq: 26 Nov 2019 15:16

Instrument :

MSVOA_D

ClientSampleId :


B-1-(0.5-1.0)

Tot Ion:168 Resp: 754863

Ion Ratio Lower Upper




168 100

99 48.7 41.6 62.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VDC



#33

1,2-Dichloroethane-d4

Concen: 50.500 ug/l

RT: 8.34 min Scan# 1091

Delta R.T. 0.00 min

Lab File: VD064384.D


Acq: 26 Nov 2019 15:16

Tgt Ion: 65 Resp: 313521

Ion Ratio Lower Upper



65 100


67 50.8 0.0 104.4



Abundance Ion 64.90 (64.60 to 65.60): VDC

Ion 66.90 (66.60 to 67.60): VDC









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.87 min Scan# 1181
 Delta R.T. 0.00 min
 Lab File: VD064384.D
 Acq: 26 Nov 2019 15:16

Instrument :
 MSVOA_D
 ClientSampleId :
 B-1-(0.5-1.0)

Tot Ion:114 Resp: 1120076
 Ion Ratio Lower Upper
 114 100
 63 19.0 0.0 40.0
 88 14.8 0.0 30.2




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VDC
 Ion 87.90 (87.60 to 88.60): VDC






#35
 Dibromofluoromethane
 Concen: 52.084 ug/l
 RT: 7.92 min Scan# 1020
 Delta R.T. 0.00 min
 Lab File: VD064384.D
 Acq: 26 Nov 2019 15:16

Tgt Ion:113 Resp: 334873
 Ion Ratio Lower Upper
 113 100
 111 101.2 81.5 122.3
 192 23.2 18.8 28.2



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 53.713 ug/l

RT: 10.34 min Scan# 1432

Delta R.T. 0.00 min

Lab File: VD064384.D

Acq: 26 Nov 2019 15:16

Instrument :

MSVOA_D

ClientSampleId :


B-1-(0.5-1.0)

Tot Ion: 98 Resp: 1302784

Ion Ratio Lower Upper




98 100

100 65.5 52.6 78.8



Abundance Ion 98.00 (97.70 to 98.70): VDC

Ion 100.00 (99.70 to 100.70): VDC



#62

4-Bromofluorobenzene

Concen: 48.007 ug/l

RT: 12.64 min Scan# 1822

Delta R.T. 0.00 min

Lab File: VD064384.D

Acq: 26 Nov 2019 15:16


Tgt Ion: 95 Resp: 403239

Ion Ratio Lower Upper

95 100

174 95.0 0.0 183.6



176 92.6 0.0 180.2




Abundance Ion 94.90 (94.60 to 95.60): VDC

Ion 173.80 (173.50 to 174.50): VDC

Ion 175.80 (175.50 to 176.50): VDC









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.65 min Scan# 1654
Delta R.T. 0.00 min
Lab File: VD064384.D
Acq: 26 Nov 2019 15:16

Instrument :
MSVOA_D
ClientSampleId :
B-1-(0.5-1.0)

Tot Ion	Ratio	Lower	Upper
117	100		
82	51.3	42.8	64.2
119	31.3	25.6	38.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.58 min Scan# 1983
Delta R.T. 0.01 min
Lab File: VD064384.D
Acq: 26 Nov 2019 15:16

Tgt Ion	Ratio	Lower	Upper
152	100		
115	54.3	27.4	82.2
150	156.8	0.0	351.6

