


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD112719\
 Data File : VD064393.D
 Acq On : 27 Nov 2019 12:57
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D112719S.M
 Title : SW846 8260
 Last Update : Fri Nov 29 07:18:03 2019



AutoFind: Scans 1821, 1822, 1823; Background Corrected with Scan 1813

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.1	20917	PASS
75	95	30	60	52.5	54653	PASS
95	95	100	100	100.0	104029	PASS
96	95	5	9	7.3	7608	PASS
173	174	0.00	2	0.8	851	PASS
174	95	50	100	97.2	101130	PASS
175	174	5	9	7.9	8032	PASS
176	174	95	101	96.5	97629	PASS
177	176	5	9	7.5	7366	PASS