



Methyl Iodide

Response Ratio



Response = 6.434e-001 * Amt - 5.869e-002
Coef of Det (r^2)= 0.998745 Curve Fit: Linear
Method Name: Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Calibration Table Last Updated: Mon Nov 29 17:54:09 2021


Tert butylalcohol



$R = 9.392e-004 A^2 + 1.984e-002 A + 6.588e-002$
 Coef of Det (r^2) = 0.993532 Curve Fit: Quadratic
 Method Name: Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Calibration Table Last Updated: Mon Nov 29 17:54:09 2021

Acetone

Response Ratio



$$\text{Response} = 1.777\text{e-}001 * \text{Amt} - 1.094\text{e-}001$$


Coef of Det (r^2) = 0.992140 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M

Calibration Table Last Updated: Mon Nov 29 17:54:09 2021

Methylene Chloride

Response Ratio



$$\text{Response} = 5.426\text{e-}001 * \text{Amt} + 4.945\text{e-}002$$


Coef of Det (r^2) = 0.999341 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M

Calibration Table Last Updated: Mon Nov 29 17:54:09 2021

Vinyl Acetate

Response Ratio



$$\text{Response} = 1.072\text{e}+000 * \text{Amt} - 5.835\text{e}-001$$


Coef of Det (r^2) = 0.995073 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M

Calibration Table Last Updated: Mon Nov 29 17:54:09 2021

2-Butanone

Response Ratio




$$\text{Response} = 2.030\text{e-}001 * \text{Amt} - 1.151\text{e-}001$$

Coef of Det (r^2) = 0.991593 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M

Calibration Table Last Updated: Mon Nov 29 17:54:09 2021


Methacrylonitrile



$R = 3.703e-002 A^2 + 4.652e-002 A + 1.261e-002$
Coef of Det (r^2) = 0.998300 Curve Fit: Quadratic
Method Name: Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Calibration Table Last Updated: Mon Nov 29 17:54:09 2021

Isopropyl Acetate

Response Ratio



$$\text{Response} = 4.960\text{e-}001 * \text{Amt} - 5.181\text{e-}002$$


Coef of Det (r^2) = 0.993769 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M

Calibration Table Last Updated: Mon Nov 29 17:54:09 2021

4-Methyl-2-Pentanone

Response Ratio



$$\text{Response} = 2.594\text{e-}001 * \text{Amt} - 1.590\text{e-}001$$


Coef of Det (r^2) = 0.989097 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M

Calibration Table Last Updated: Mon Nov 29 17:54:09 2021

Ethyl methacrylate

Response Ratio



$$\text{Response} = 3.714\text{e-}001 * \text{Amt} - 4.435\text{e-}002$$


Coef of Det (r^2) = 0.992706 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M

Calibration Table Last Updated: Mon Nov 29 17:54:09 2021

2-Hexanone

Response Ratio



$$\text{Response} = 1.990\text{e-}001 * \text{Amt} - 1.376\text{e-}001$$


Coef of Det (r^2) = 0.989775 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M

Calibration Table Last Updated: Mon Nov 29 17:54:09 2021

N-amyl acetate

Response Ratio




$$\text{Response} = 1.119\text{e}+000 * \text{Amt} - 1.276\text{e}-001$$

Coef of Det (r^2) = 0.992460 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M

Calibration Table Last Updated: Mon Nov 29 17:54:09 2021

trans-1,4-Dichloro-2-butene



Response = $2.303 \times 10^{-1} \times \text{Amt} - 2.659 \times 10^{-2}$
Coef of Det (r^2) = 0.990868 Curve Fit: Linear
Method Name: Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Calibration Table Last Updated: Mon Nov 29 17:54:09 2021