



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD112921\
Data File : VD071070.D
Acq On : 29 Nov 2021 12:18
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Title : SW846 8260
Last Update : Mon Nov 29 17:54:09 2021



AutoFind: Scans 1819, 1820, 1821; Background Corrected with Scan 1810

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	22.3	30832	PASS
75	95	30	60	54.9	76029	PASS
95	95	100	100	100.0	138368	PASS
96	95	5	9	6.7	9281	PASS
173	174	0.00	2	1.7	1756	PASS
174	95	50	100	76.3	105613	PASS
175	174	5	9	8.0	8482	PASS
176	174	95	101	96.3	101683	PASS
177	176	5	9	8.4	8517	PASS