

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD113021\
 Data File : VD071081.D
 Acq On : 30 Nov 2021 11:43
 Operator : VA/SY
 Sample : VD1130SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1130SBL01

Quant Time: Dec 01 04:19:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	292077	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	502389	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	447738	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	194246	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	165984	48.432	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.860%
35) Dibromofluoromethane	7.920	113	156482	48.913	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.820%
50) Toluene-d8	10.338	98	574677	47.204	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	94.400%
62) 4-Bromofluorobenzene	12.626	95	193624	45.881	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	91.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD113021\
Data File : VD071081.D
Acq On : 30 Nov 2021 11:43
Operator : VA/SY
Sample : VD1130SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1130SBL01

Quant Time: Dec 01 04:19:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

