

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_D\Data\VD113021\  
 Data File : VD071083.D  
 Acq On : 30 Nov 2021 12:50  
 Operator : VA/SY  
 Sample : VD1130SBS01  
 Misc : 5.00G/5.00ml/MSVOA\_D/Soil  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 VD1130SBS01

Manual Integrations  
 APPROVED

Reviewed By : Semsettin  
 Yesilyurt

Quant Time: Dec 01 03:32:09 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_D\Method\82D112921S.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Nov 29 17:54:09 2021  
 Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 7.979  | 168  | 307787   | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 8.861  | 114  | 511976   | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 11.637 | 117  | 462992   | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 13.567 | 152  | 211870   | 50.000 | ug/l  | 0.00     |

12/01/2021  
 Supervised By : Mahesh  
 Adoda

|                             |        |       |          |          |      |         |
|-----------------------------|--------|-------|----------|----------|------|---------|
| System Monitoring Compounds |        |       |          |          |      |         |
| 33) 1,2-Dichloroethane-d4   | 8.332  | 65    | 154349   | 42.738   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 50 - 163 | Recovery | =    | 85.480% |
| 35) Dibromofluoromethane    | 7.914  | 113   | 144313   | 44.265   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 54 - 147 | Recovery | =    | 88.520% |
| 50) Toluene-d8              | 10.337 | 98    | 533346   | 42.988   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 49 - 140 | Recovery | =    | 85.980% |
| 62) 4-Bromofluorobenzene    | 12.626 | 95    | 188708   | 43.879   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 25 - 144 | Recovery | =    | 87.760% |

12/01/2021

|                              |       |     |        |         |        |        |
|------------------------------|-------|-----|--------|---------|--------|--------|
| Target Compounds             |       |     |        |         |        | Qvalue |
| 2) Dichlorodifluoromethane   | 1.991 | 85  | 49824  | 18.983  | ug/l   | 95     |
| 3) Chloromethane             | 2.214 | 50  | 66306  | 19.095  | ug/l   | 95     |
| 4) Vinyl Chloride            | 2.356 | 62  | 71554  | 18.470  | ug/l   | 99     |
| 5) Bromomethane              | 2.773 | 94  | 54081  | 19.365  | ug/l   | 99     |
| 6) Chloroethane              | 2.932 | 64  | 49233  | 18.477  | ug/l   | 92     |
| 7) Trichlorofluoromethane    | 3.279 | 101 | 94998  | 18.102  | ug/l   | 98     |
| 8) Diethyl Ether             | 3.708 | 74  | 24413  | 17.159  | ug/l   | 88     |
| 9) 1,1,2-Trichlorotrifluo... | 4.108 | 101 | 54260  | 17.525  | ug/l   | 96     |
| 10) Methyl Iodide            | 4.314 | 142 | 47474  | 16.547  | ug/l   | 96     |
| 11) Tert butyl alcohol       | 5.208 | 59  | 41280  | 150.522 | ug/l # | 80     |
| 12) 1,1-Dichloroethene       | 4.079 | 96  | 49632  | 17.370  | ug/l   | 89     |
| 13) Acrolein                 | 3.926 | 56  | 25781  | 104.461 | ug/l   | 99     |
| 14) Allyl chloride           | 4.720 | 41  | 92713  | 17.083  | ug/l # | 82     |
| 15) Acrylonitrile            | 5.432 | 53  | 59330  | 85.747  | ug/l   | 97     |
| 16) Acetone                  | 4.155 | 43  | 57062  | 82.968  | ug/l   | 90     |
| 17) Carbon Disulfide         | 4.414 | 76  | 151754 | 18.140  | ug/l   | 98     |
| 18) Methyl Acetate           | 4.720 | 43  | 47444  | 17.805  | ug/l   | 93     |
| 19) Methyl tert-butyl Ether  | 5.485 | 73  | 117676 | 16.805  | ug/l   | 99     |
| 20) Methylene Chloride       | 4.973 | 84  | 67269  | 15.584  | ug/l   | 89     |
| 21) trans-1,2-Dichloroethene | 5.479 | 96  | 55062  | 17.119  | ug/l   | 82     |
| 22) Diisopropyl ether        | 6.361 | 45  | 181828 | 16.722  | ug/l   | 93     |
| 23) Vinyl Acetate            | 6.308 | 43  | 405179 | 88.622  | ug/l   | 98     |
| 24) 1,1-Dichloroethane       | 6.267 | 63  | 107063 | 17.118  | ug/l   | 97     |
| 25) 2-Butanone               | 7.208 | 43  | 77205  | 90.157  | ug/l # | 83     |
| 26) 2,2-Dichloropropane      | 7.214 | 77  | 94412  | 16.601  | ug/l   | 95     |
| 27) cis-1,2-Dichloroethene   | 7.214 | 96  | 63084  | 17.337  | ug/l   | 91     |
| 28) Bromochloromethane       | 7.543 | 49  | 55572  | 20.457  | ug/l   | 90     |
| 29) Tetrahydrofuran          | 7.561 | 42  | 46404  | 80.917  | ug/l   | 95     |
| 30) Chloroform               | 7.708 | 83  | 107063 | 17.057  | ug/l   | 99     |
| 31) Cyclohexane              | 7.985 | 56  | 105515 | 17.176  | ug/l # | 93     |
| 32) 1,1,1-Trichloroethane    | 7.902 | 97  | 96206  | 17.072  | ug/l   | 94     |
| 36) 1,1-Dichloropropene      | 8.114 | 75  | 81376  | 17.006  | ug/l   | 98     |
| 37) Ethyl Acetate            | 7.291 | 43  | 34166  | 16.529  | ug/l   | 95     |
| 38) Carbon Tetrachloride     | 8.096 | 117 | 84114  | 17.832  | ug/l   | 97     |
| 39) Methylcyclohexane        | 9.355 | 83  | 100291 | 17.426  | ug/l   | 99     |
| 40) Benzene                  | 8.355 | 78  | 229271 | 17.639  | ug/l   | 99     |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_D\Data\VD113021\  
 Data File : VD071083.D  
 Acq On : 30 Nov 2021 12:50  
 Operator : VA/SY  
 Sample : VD1130SBS01  
 Misc : 5.00G/5.00ml/MSVOA\_D/SoIL  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 VD1130SBS01

Manual Integrations  
 APPROVED

Reviewed By : Semsettin  
 Yesilyurt

Quant Time: Dec 01 03:32:09 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_D\Method\82D112921S.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Nov 29 17:54:09 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units | Dev(Min) |     |
|-------------------------------|--------|------|----------|---------|-------|----------|-----|
| 41) Methacrylonitrile         | 7.520  | 41   | 18366    | 19.160  | ug/l  | #        | 83  |
| 42) 1,2-Dichloroethane        | 8.426  | 62   | 69351    | 17.480  | ug/l  |          | 95  |
| 43) Isopropyl Acetate         | 8.455  | 43   | 68138    | 18.639  | ug/l  |          | 91  |
| 44) Trichloroethene           | 9.108  | 130  | 60654    | 17.502  | ug/l  |          | 93  |
| 45) 1,2-Dichloropropane       | 9.385  | 63   | 59509    | 17.410  | ug/l  |          | 91  |
| 46) Dibromomethane            | 9.473  | 93   | 30328    | 17.102  | ug/l  |          | 92  |
| 47) Bromodichloromethane      | 9.661  | 83   | 77803    | 16.589  | ug/l  | #        | 88  |
| 48) Methyl methacrylate       | 9.449  | 41   | 33117    | 16.368  | ug/l  | #        | 79  |
| 49) 1,4-Dioxane               | 9.467  | 88   | 5577     | 325.784 | ug/l  | #        | 83  |
| 51) 4-Methyl-2-Pentanone      | 10.226 | 43   | 173342   | 95.904  | ug/l  |          | 91  |
| 52) Toluene                   | 10.402 | 92   | 141187   | 17.120  | ug/l  |          | 99  |
| 53) t-1,3-Dichloropropene     | 10.614 | 75   | 73362    | 17.014  | ug/l  |          | 99  |
| 54) cis-1,3-Dichloropropene   | 10.085 | 75   | 84796    | 16.555  | ug/l  | #        | 78  |
| 55) 1,1,2-Trichloroethane     | 10.796 | 97   | 39988    | 17.189  | ug/l  |          | 96  |
| 56) Ethyl methacrylate        | 10.655 | 69   | 49899    | 19.090  | ug/l  | #        | 75  |
| 57) 1,3-Dichloropropane       | 10.943 | 76   | 69042    | 16.894  | ug/l  |          | 100 |
| 58) 2-Chloroethyl Vinyl ether | 9.937  | 63   | 161977   | 102.317 | ug/l  |          | 96  |
| 59) 2-Hexanone                | 10.979 | 43   | 116722   | 91.844  | ug/l  |          | 91  |
| 60) Dibromochloromethane      | 11.137 | 129  | 47750    | 16.515  | ug/l  |          | 97  |
| 61) 1,2-Dibromoethane         | 11.243 | 107  | 39317    | 17.597  | ug/l  |          | 97  |
| 64) Tetrachloroethene         | 10.879 | 164  | 50215    | 17.084  | ug/l  |          | 89  |
| 65) Chlorobenzene             | 11.667 | 112  | 147712   | 17.195  | ug/l  |          | 98  |
| 66) 1,1,1,2-Tetrachloroethane | 11.737 | 131  | 54930    | 17.513  | ug/l  |          | 98  |
| 67) Ethyl Benzene             | 11.737 | 91   | 270507   | 16.727  | ug/l  |          | 99  |
| 68) m/p-Xylenes               | 11.849 | 106  | 205239   | 33.616  | ug/l  |          | 96  |
| 69) o-Xylene                  | 12.173 | 106  | 94992    | 16.554  | ug/l  |          | 96  |
| 70) Styrene                   | 12.190 | 104  | 160520   | 16.598  | ug/l  |          | 96  |
| 71) Bromoform                 | 12.349 | 173  | 27393    | 16.858  | ug/l  | #        | 95  |
| 73) Isopropylbenzene          | 12.473 | 105  | 261282   | 17.031  | ug/l  |          | 98  |
| 74) N-amyl acetate            | 12.284 | 43   | 59500    | 18.252  | ug/l  |          | 90  |
| 75) 1,1,2,2-Tetrachloroethane | 12.726 | 83   | 45004    | 17.814  | ug/l  |          | 98  |
| 76) 1,2,3-Trichloropropane    | 12.773 | 75   | 30991m   | 15.353  | ug/l  |          |     |
| 77) Bromobenzene              | 12.755 | 156  | 55889    | 16.819  | ug/l  |          | 99  |
| 78) n-propylbenzene           | 12.814 | 91   | 319119   | 16.704  | ug/l  |          | 96  |
| 79) 2-Chlorotoluene           | 12.896 | 91   | 183063   | 17.072  | ug/l  |          | 99  |
| 80) 1,3,5-Trimethylbenzene    | 12.949 | 105  | 215710   | 17.051  | ug/l  |          | 97  |
| 81) trans-1,4-Dichloro-2-b... | 12.520 | 75   | 12301    | 18.380  | ug/l  | #        | 76  |
| 82) 4-Chlorotoluene           | 12.996 | 91   | 185581   | 16.671  | ug/l  |          | 100 |
| 83) tert-Butylbenzene         | 13.214 | 119  | 181809   | 16.735  | ug/l  |          | 95  |
| 84) 1,2,4-Trimethylbenzene    | 13.261 | 105  | 211608   | 17.133  | ug/l  |          | 99  |
| 85) sec-Butylbenzene          | 13.396 | 105  | 279689   | 16.855  | ug/l  |          | 99  |
| 86) p-Isopropyltoluene        | 13.508 | 119  | 228399   | 16.829  | ug/l  |          | 98  |
| 87) 1,3-Dichlorobenzene       | 13.508 | 146  | 112815   | 17.055  | ug/l  |          | 98  |
| 88) 1,4-Dichlorobenzene       | 13.584 | 146  | 112906   | 17.228  | ug/l  |          | 98  |
| 89) n-Butylbenzene            | 13.831 | 91   | 224376   | 16.936  | ug/l  |          | 99  |
| 90) Hexachloroethane          | 14.102 | 117  | 43639    | 16.962  | ug/l  |          | 92  |
| 91) 1,2-Dichlorobenzene       | 13.878 | 146  | 96574    | 17.168  | ug/l  |          | 99  |
| 92) 1,2-Dibromo-3-Chloropr... | 14.496 | 75   | 7060     | 17.160  | ug/l  |          | 98  |
| 93) 1,2,4-Trichlorobenzene    | 15.149 | 180  | 59884    | 17.067  | ug/l  |          | 98  |
| 94) Hexachlorobutadiene       | 15.249 | 225  | 35434    | 17.848  | ug/l  |          | 97  |
| 95) Naphthalene               | 15.384 | 128  | 101231   | 16.366  | ug/l  |          | 99  |
| 96) 1,2,3-Trichlorobenzene    | 15.578 | 180  | 49907    | 16.746  | ug/l  |          | 95  |

12/01/2021  
 Supervised By : Mahesh  
 Dadoda

12/01/2021

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_D\Data\VD113021\  
 Data File : VD071083.D  
 Acq On : 30 Nov 2021 12:50  
 Operator : VA/SY  
 Sample : VD1130SBSD01  
 Misc : 5.00G/5.00ml/MSVOA\_D/SOIL  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 VD1130SBSD01

Manual Integrations  
 APPROVED

Reviewed By :Semsettin  
 Yesilyurt

Quant Time: Dec 01 03:32:09 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_D\Method\82D112921S.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Nov 29 17:54:09 2021  
 Response via : Initial Calibration


| Compound | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------|------|------|----------|------|-------|----------|
|----------|------|------|----------|------|-------|----------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

12/01/2021  
 Supervised By :Mahesh  
 Dadoda


12/01/2021





Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 7.979 min Scan# 10  
Delta R.T. -0.006 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01



Tgt Ion:168 Resp: 30778  
Ion Ratio Lower Upper  
168 100  
99 62.4 46.6 69.8


Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt






12/01/2021  
Supervised By :Mahesh  
Dadoda


12/01/2021



Dichlorodifluoromethane  
Concen: 18.983 ug/l  
RT: 1.991 min Scan# 12  
Delta R.T. -0.000 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50


Tgt Ion: 85 Resp: 49824  
Ion Ratio Lower Upper  
85 100  
87 33.4 15.2 45.6





#3  
Chloromethane  
Concen: 19.095 ug/l  
RT: 2.214 min Scan# 50  
Delta R.T. -0.001 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01



Tgt Ion: 50 Resp: 66300  
Ion Ratio Lower Upper  
50 100  
52 35.1 25.8 38.8


Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt




12/01/2021  
Supervised By :Mahesh  
Dadoda



12/01/2021




#4  
Vinyl Chloride  
Concen: 18.470 ug/l  
RT: 2.356 min Scan# 74  
Delta R.T. -0.000 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50




Tgt Ion: 62 Resp: 71554  
Ion Ratio Lower Upper  
62 100  
64 33.1 26.0 39.0





#5  
Bromomethane  
Concen: 19.365 ug/l  
RT: 2.773 min Scan# 146  
Delta R.T. -0.006 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Instrument :  
MSVOA\_D  
Client Sample Id :  
VD1130SBSD01



Tgt Ion: 94 Resp: 5408  
Ion Ratio Lower Upper  
94 100  
96 92.2 74.8 112.2


Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt






12/01/2021  
Supervised By :Mahesh  
Dadoda


12/01/2021



#6  
Chloroethane  
Concen: 18.477 ug/l  
RT: 2.932 min Scan# 172  
Delta R.T. -0.000 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Tgt Ion: 64 Resp: 49233  
Ion Ratio Lower Upper  
64 100  
66 31.7 29.0 43.4







#7  
Trichlorofluoromethane  
Concen: 18.102 ug/l  
RT: 3.279 min Scan# 21  
Delta R.T. 0.000 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBSD01

Tgt Ion:101 Resp: 94998  
Ion Ratio Lower Upper  
101 100  
103 63.3 49.7 74.5



Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt






12/01/2021  
Supervised By :Mahesh  
Dadoda


12/01/2021



#8  
Diethyl Ether  
Concen: 17.159 ug/l  
RT: 3.708 min Scan# 304  
Delta R.T. -0.007 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Tgt Ion: 74 Resp: 24413  
Ion Ratio Lower Upper  
74 100  
45 111.0 49.6 149.0







#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 17.525 ug/l  
RT: 4.108 min Scan# 31  
Delta R.T. 0.005 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01

Tgt Ion:101 Resp: 54260  
Ion Ratio Lower Upper  
101 100  
85 45.2 34.8 52.2  
151 71.0 60.4 90.6



Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt






12/01/2021  
Supervised By :Mahesh  
Dadoda


12/01/2021



#10  
Methyl Iodide  
Concen: 16.547 ug/l  
RT: 4.314 min Scan# 407  
Delta R.T. 0.011 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50


Tgt Ion:142 Resp: 47474  
Ion Ratio Lower Upper  
142 100  
127 46.7 35.0 52.4  
141 14.7 12.0 18.0





#11  
Tert butyl alcohol  
Concen: 150.522 ug/l  
RT: 5.208 min Scan# 511  
Delta R.T. -0.012 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01



Tgt Ion: 59 Resp: 41280  
Ion Ratio Lower Upper  
59 100  
57 3.4 9.0 13.4


Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt




12/01/2021  
Supervised By :Mahesh  
Dadoda



12/01/2021




#12  
1,1-Dichloroethene  
Concen: 17.370 ug/l  
RT: 4.079 min Scan# 367  
Delta R.T. 0.006 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50




Tgt Ion: 96 Resp: 49632  
Ion Ratio Lower Upper  
96 100  
61 170.5 120.3 180.5  
98 68.2 54.4 81.6





#13  
Acrolein  
Concen: 104.461 ug/l  
RT: 3.926 min Scan# 341  
Delta R.T. 0.000 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBSD01



Tgt Ion: 56 Resp: 2578  
Ion Ratio Lower Upper  
56 100  
55 73.8 58.6 87.8


Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt




12/01/2021  
Supervised By :Mahesh  
Dadoda



12/01/2021




#14  
Allyl chloride  
Concen: 17.083 ug/l  
RT: 4.720 min Scan# 476  
Delta R.T. -0.001 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50




Tgt Ion: 41 Resp: 92713  
Ion Ratio Lower Upper  
41 100  
39 75.8 45.6 68.4#  
76 33.5 30.2 45.4





#15  
Acrylonitrile  
Concen: 85.747 ug/l  
RT: 5.432 min Scan# 59330  
Delta R.T. 0.012 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01



Tgt Ion: 53 Resp: 59330  
Ion Ratio Lower Upper  
53 100  
52 77.9 63.8 95.8  
51 38.7 28.5 42.7


Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt




12/01/2021  
Supervised By :Mahesh  
Dadoda



12/01/2021




#16  
Acetone  
Concen: 82.968 ug/l  
RT: 4.155 min Scan# 380  
Delta R.T. -0.001 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50




Tgt Ion: 43 Resp: 57062  
Ion Ratio Lower Upper  
43 100  
58 28.9 27.8 41.8





#17  
Carbon Disulfide  
Concen: 18.140 ug/l  
RT: 4.414 min Scan# 41  
Delta R.T. -0.001 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBSD01



Tgt Ion: 76 Resp: 151754  
Ion Ratio Lower Upper  
76 100  
78 9.1 7.8 11.8


Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt






12/01/2021  
Supervised By :Mahesh  
Dadoda


12/01/2021



#18  
Methyl Acetate  
Concen: 17.805 ug/l  
RT: 4.720 min Scan# 476  
Delta R.T. 0.005 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50


Tgt Ion: 43 Resp: 47444  
Ion Ratio Lower Upper  
43 100  
74 20.9 19.3 28.9





#19  
Methyl tert-butyl Ether  
Concen: 16.805 ug/l  
RT: 5.485 min Scan# 605  
Delta R.T. 0.006 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50


Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01



Tgt Ion: 73 Resp: 117670  
Ion Ratio Lower Upper  
73 100  
57 23.0 18.9 28.3


Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt






12/01/2021  
Supervised By :Mahesh  
Dadoda


12/01/2021



#20  
Methylene Chloride  
Concen: 15.584 ug/l  
RT: 4.973 min Scan# 519  
Delta R.T. 0.005 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50


Tgt Ion: 84 Resp: 67269  
Ion Ratio Lower Upper  
84 100  
49 137.2 95.8 143.6  
51 43.7 30.9 46.3  
86 60.9 50.3 75.5





#21  
trans-1,2-Dichloroethene  
Concen: 17.119 ug/l  
RT: 5.479 min Scan# 605  
Delta R.T. -0.000 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01



Tgt Ion: 96 Resp: 55062  
Ion Ratio Lower Upper  
96 100  
61 151.6 102.1 153.1  
98 54.2 50.1 75.1


Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt




12/01/2021  
Supervised By :Mahesh  
Dadoda



12/01/2021




#22  
Diisopropyl ether  
Concen: 16.722 ug/l  
RT: 6.361 min Scan# 755  
Delta R.T. -0.006 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Tgt Ion: 45 Resp: 181828  
Ion Ratio Lower Upper  
45 100  
43 56.5 39.6 59.4  
87 24.3 21.2 31.8  
59 10.2 9.0 13.6







#23  
 Vinyl Acetate  
 Concen: 88.622 ug/l  
 RT: 6.308 min Scan# 741  
 Delta R.T. -0.001 min  
 Lab File: VD071083.D  
 Acq: 30 Nov 2021 12:50

Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 VD1130SBSD01

Tgt Ion: 43 Resp: 405179  
 Ion Ratio Lower Upper  
 43 100  
 86 9.5 8.2 12.4



Manual Integrations  
 APPROVED

Reviewed By :Semsettin  
 Yesilyurt






12/01/2021  
 Supervised By :Mahesh  
 Dadoda


12/01/2021



#24  
 1,1-Dichloroethane  
 Concen: 17.118 ug/l  
 RT: 6.267 min Scan# 739  
 Delta R.T. 0.005 min  
 Lab File: VD071083.D  
 Acq: 30 Nov 2021 12:50


Tgt Ion: 63 Resp: 107063  
 Ion Ratio Lower Upper  
 63 100  
 98 8.1 3.5 10.6  
 100 5.3 2.1 6.3





#25  
2-Butanone  
Concen: 90.157 ug/l  
RT: 7.208 min Scan# 899  
Delta R.T. -0.007 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBSD01



Tgt Ion: 43 Resp: 7720  
Ion Ratio Lower Upper  
43 100  
72 20.6 23.8 35.6


Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt






12/01/2021  
Supervised By :Mahesh  
Dadoda


12/01/2021



#26  
2,2-Dichloropropane  
Concen: 16.601 ug/l  
RT: 7.214 min Scan# 900  
Delta R.T. 0.005 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50


Tgt Ion: 77 Resp: 94412  
Ion Ratio Lower Upper  
77 100  
97 21.2 11.9 35.5





#27  
cis-1,2-Dichloroethene  
Concen: 17.337 ug/l  
RT: 7.214 min Scan# 900  
Delta R.T. -0.001 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01



Tgt Ion: 96 Resp: 63084  
Ion Ratio Lower Upper  
96 100  
61 159.5 0.0 288.6  
98 62.4 0.0 129.2


Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt






12/01/2021  
Supervised By :Mahesh  
Dadoda


12/01/2021



#28  
Bromochloromethane  
Concen: 20.457 ug/l  
RT: 7.543 min Scan# 956  
Delta R.T. -0.001 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50


Tgt Ion: 49 Resp: 55572  
Ion Ratio Lower Upper  
49 100  
129 1.3 0.0 4.0  
130 66.2 60.2 90.2





#29  
Tetrahydrofuran  
Concen: 80.917 ug/l  
RT: 7.561 min Scan# 91  
Delta R.T. -0.001 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01



Tgt Ion: 42 Resp: 46404  
Ion Ratio Lower Upper  
42 100  
72 39.8 33.7 50.5  
71 37.0 32.7 49.1


Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt




12/01/2021  
Supervised By :Mahesh  
Dadoda



12/01/2021




#30  
Chloroform  
Concen: 17.057 ug/l  
RT: 7.708 min Scan# 984  
Delta R.T. -0.001 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50




Tgt Ion: 83 Resp: 107063  
Ion Ratio Lower Upper  
83 100  
85 64.2 51.0 76.6





#31  
Cyclohexane  
Concen: 17.176 ug/l  
RT: 7.985 min Scan# 1031  
Delta R.T. -0.000 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBSD01



Tgt Ion: 56 Resp: 10551  
Ion Ratio Lower Upper  
56 100  
69 39.9 25.9 38.9  
84 76.8 64.3 96.5


Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt




12/01/2021  
Supervised By :Mahesh  
Dadoda



12/01/2021




#32  
1,1,1-Trichloroethane  
Concen: 17.072 ug/l  
RT: 7.902 min Scan# 1017  
Delta R.T. -0.001 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Tgt Ion: 97 Resp: 96206  
Ion Ratio Lower Upper  
97 100  
99 61.4 50.3 75.5  
61 48.4 33.0 49.4







#33  
1,2-Dichloroethane-d4  
Concen: 42.738 ug/l  
RT: 8.332 min Scan# 1089  
Delta R.T. 0.006 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01

Tgt Ion: 65 Resp: 154349  
Ion Ratio Lower Upper  
65 100  
67 51.8 0.0 112.2



Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt






12/01/2021  
Supervised By :Mahesh  
Dadoda


12/01/2021



#34  
1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 8.861 min Scan# 1180  
Delta R.T. 0.000 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50


Tgt Ion:114 Resp: 511976  
Ion Ratio Lower Upper  
114 100  
63 21.7 0.0 41.8  
88 15.6 0.0 32.8





#35  
Dibromofluoromethane  
Concen: 44.265 ug/l  
RT: 7.914 min Scan# 1019  
Delta R.T. -0.006 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01



Tgt Ion: 113 Resp: 14431  
Ion Ratio Lower Upper  
113 100  
111 102.2 82.0 123.0  
192 19.0 15.2 22.8


Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt




12/01/2021  
Supervised By :Mahesh  
Dadoda



12/01/2021




#36  
1,1-Dichloropropene  
Concen: 17.006 ug/l  
RT: 8.114 min Scan# 1053  
Delta R.T. 0.000 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Tgt Ion: 75 Resp: 81376  
Ion Ratio Lower Upper  
75 100  
110 33.9 17.8 53.4  
77 31.1 24.8 37.2





#37

Ethyl Acetate

Concen: 16.529 ug/l

RT: 7.291 min Scan# 912

Delta R.T. 0.006 min

Lab File: VD071083.D

Acq: 30 Nov 2021 12:50

Instrument :

MSVOA\_D

ClientSampleId :



VD1130SBS01

Tgt Ion: 43 Resp: 3416

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 43  | 100   |       |       |
| 61  | 13.7  | 13.0  | 19.6  |
| 70  | 9.2   | 8.7   | 13.1  |



Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt



12/01/2021  
Supervised By :Mahesh  
Dadoda

12/01/2021



#38

Carbon Tetrachloride

Concen: 17.832 ug/l

RT: 8.096 min Scan# 1050




Delta R.T. -0.001 min


Lab File: VD071083.D

Acq: 30 Nov 2021 12:50

Tgt Ion:117 Resp: 84114


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 117 | 100   |       |       |
| 119 | 91.8  | 75.8  | 113.6 |
| 121 | 30.5  | 23.6  | 35.4  |





#39  
Methylcyclohexane  
Concen: 17.426 ug/l  
RT: 9.355 min Scan# 11094  
Delta R.T. -0.001 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01



Tgt Ion: 83 Resp: 100291  
Ion Ratio Lower Upper  
83 100  
55 81.7 65.8 98.8  
98 48.9 38.7 58.1


Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt




12/01/2021  
Supervised By :Mahesh  
Dadoda



12/01/2021




#40  
Benzene  
Concen: 17.639 ug/l  
RT: 8.355 min Scan# 1094  
Delta R.T. 0.005 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50




Tgt Ion: 78 Resp: 229271  
Ion Ratio Lower Upper  
78 100  
77 23.2 19.1 28.7





#41  
Methacrylonitrile  
Concen: 19.160 ug/l  
RT: 7.520 min Scan# 91  
Delta R.T. -0.000 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBSD01



Tgt Ion: 41 Resp: 18360  
Ion Ratio Lower Upper  
41 100  
39 64.5 60.2 90.4  
67 68.9 74.8 112.2  
52 34.2 28.8 43.2


Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt






12/01/2021  
Supervised By :Mahesh  
Dadoda


12/01/2021



#42  
1,2-Dichloroethane  
Concen: 17.480 ug/l  
RT: 8.426 min Scan# 1106  
Delta R.T. 0.006 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50


Tgt Ion: 62 Resp: 69351  
Ion Ratio Lower Upper  
62 100  
98 8.9 0.0 21.8





#43  
Isopropyl Acetate  
Concen: 18.639 ug/l  
RT: 8.455 min Scan# 1110  
Delta R.T. 0.005 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01



Tgt Ion: 43 Resp: 68138  
Ion Ratio Lower Upper  
43 100  
61 28.8 28.2 42.4  
87 12.0 10.7 16.1


Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt






12/01/2021  
Supervised By :Mahesh  
Dadoda


12/01/2021



#44  
Trichloroethene  
Concen: 17.502 ug/l  
RT: 9.108 min Scan# 1222  
Delta R.T. -0.001 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Tgt Ion:130 Resp: 60654  
Ion Ratio Lower Upper  
130 100  
95 103.8 0.0 193.2







#45  
1,2-Dichloropropane  
Concen: 17.410 ug/l  
RT: 9.385 min Scan# 1269  
Delta R.T. -0.000 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01

Tgt Ion: 63 Resp: 59509  
Ion Ratio Lower Upper  
63 100  
65 35.3 24.2 36.2



Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt






12/01/2021  
Supervised By :Mahesh  
Dadoda


12/01/2021



#46  
Dibromomethane  
Concen: 17.102 ug/l  
RT: 9.473 min Scan# 1284  
Delta R.T. -0.000 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Tgt Ion: 93 Resp: 30328  
Ion Ratio Lower Upper  
93 100  
95 85.5 66.1 99.1  
174 87.1 78.6 117.8







#47  
Bromodichloromethane  
Concen: 16.589 ug/l  
RT: 9.661 min Scan# 11  
Delta R.T. 0.000 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01

Tgt Ion: 83 Resp: 7780  
Ion Ratio Lower Upper  
83 100  
85 72.1 50.0 75.0  
127 11.0 7.3 10.9



Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt






12/01/2021  
Supervised By :Mahesh  
Dadoda


12/01/2021



#48  
Methyl methacrylate  
Concen: 16.368 ug/l  
RT: 9.449 min Scan# 1280  
Delta R.T. -0.007 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Tgt Ion: 41 Resp: 33117  
Ion Ratio Lower Upper  
41 100  
69 70.7 69.8 104.6  
39 63.0 35.6 53.4





#49

1,4-Dioxane

Concen: 325.784 ug/l

RT: 9.467 min Scan# 1282

Delta R.T. 0.006 min

Lab File: VD071083.D

Acq: 30 Nov 2021 12:50

Instrument :

MSVOA\_D

ClientSampleId :



VD1130SBS01

Tgt Ion: 88 Resp: 557

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 88  | 100   |       |       |
| 43  | 51.2  | 28.9  | 43.3  |
| 58  | 80.9  | 56.3  | 84.5  |



Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt



12/01/2021  
Supervised By :Mahesh  
Dadoda

12/01/2021



#50

Toluene-d8

Concen: 42.988 ug/l

RT: 10.337 min Scan# 1431




Delta R.T. -0.001 min


Lab File: VD071083.D

Acq: 30 Nov 2021 12:50

Tgt Ion: 98 Resp: 533346

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 98  | 100   |       |       |
| 100 | 65.5  | 51.8  | 77.6  |







#51  
4-Methyl-2-Pentanone  
Concen: 95.904 ug/l  
RT: 10.226 min Scan# 1412  
Delta R.T. -0.000 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01

Tgt Ion: 43 Resp: 17334  
Ion Ratio Lower Upper  
43 100  
58 35.7 33.1 49.7



Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt






12/01/2021  
Supervised By :Mahesh  
Dadoda


12/01/2021



#52  
Toluene  
Concen: 17.120 ug/l  
RT: 10.402 min Scan# 1442  
Delta R.T. -0.001 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Tgt Ion: 92 Resp: 141187  
Ion Ratio Lower Upper  
92 100  
91 176.9 140.7 211.1







#53  
 t-1,3-Dichloropropene  
 Concen: 17.014 ug/l  
 RT: 10.614 min Scan# 1478  
 Delta R.T. -0.000 min  
 Lab File: VD071083.D  
 Acq: 30 Nov 2021 12:50

Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 VD1130SBS01

Tgt Ion: 75 Resp: 73362  
 Ion Ratio Lower Upper  
 75 100  
 77 30.0 24.5 36.7



Manual Integrations  
 APPROVED

Reviewed By :Semsettin  
 Yesilyurt






12/01/2021  
 Supervised By :Mahesh  
 Dadoda


12/01/2021



#54  
 cis-1,3-Dichloropropene  
 Concen: 16.555 ug/l  
 RT: 10.085 min Scan# 1388  
 Delta R.T. -0.000 min  
 Lab File: VD071083.D  
 Acq: 30 Nov 2021 12:50


Tgt Ion: 75 Resp: 84796  
 Ion Ratio Lower Upper  
 75 100  
 77 30.8 25.0 37.6  
 39 63.1 31.6 47.4#





#55  
1,1,2-Trichloroethane  
Concen: 17.189 ug/l  
RT: 10.796 min Scan# 1509  
Delta R.T. -0.001 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Instrument : MSVOA\_D  
ClientSampleId : VD1130SBS01



Tgt Ion: 97 Resp: 39988  
Ion Ratio Lower Upper  
97 100  
83 80.3 65.2 97.8  
85 55.7 43.4 65.0  
99 57.9 52.1 78.1


Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt




12/01/2021  
Supervised By :Mahesh  
Dadoda



12/01/2021




#56  
Ethyl methacrylate  
Concen: 19.090 ug/l  
RT: 10.655 min Scan# 1485  
Delta R.T. -0.001 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Tgt Ion: 69 Resp: 49899  
Ion Ratio Lower Upper  
69 100  
41 76.2 50.2 75.2#  
39 46.7 19.8 29.8#







#57  
1,3-Dichloropropane  
Concen: 16.894 ug/l  
RT: 10.943 min Scan# 1534  
Delta R.T. -0.001 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01

Tgt Ion: 76 Resp: 6904  
Ion Ratio Lower Upper  
76 100  
78 32.2 25.8 38.6



Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt






12/01/2021  
Supervised By :Mahesh  
Dadoda


12/01/2021



#58  
2-Chloroethyl Vinyl ether  
Concen: 102.317 ug/l  
RT: 9.937 min Scan# 1363  
Delta R.T. -0.001 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50


Tgt Ion: 63 Resp: 161977  
Ion Ratio Lower Upper  
63 100  
106 28.1 21.0 31.4





#59  
2-Hexanone  
Concen: 91.844 ug/l  
RT: 10.979 min Scan# 11  
Delta R.T. -0.000 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBSD01



Tgt Ion: 43 Resp: 11672  
Ion Ratio Lower Upper  
43 100  
58 51.1 28.7 86.1


Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt






12/01/2021  
Supervised By :Mahesh  
Dadoda


12/01/2021



#60  
Dibromochloromethane  
Concen: 16.515 ug/l  
RT: 11.137 min Scan# 1567  
Delta R.T. -0.001 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Tgt Ion:129 Resp: 47750  
Ion Ratio Lower Upper  
129 100  
127 80.1 38.9 116.6







#61  
1,2-Dibromoethane  
Concen: 17.597 ug/l  
RT: 11.243 min Scan# 11  
Delta R.T. -0.001 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01

Tgt Ion:107 Resp: 3931  
Ion Ratio Lower Upper  
107 100  
109 90.1 74.6 112.0



Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt






12/01/2021  
Supervised By :Mahesh  
Dadoda


12/01/2021



#62  
4-Bromofluorobenzene  
Concen: 43.879 ug/l  
RT: 12.626 min Scan# 1820  
Delta R.T. -0.000 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50


Tgt Ion: 95 Resp: 188708  
Ion Ratio Lower Upper  
95 100  
174 77.0 0.0 159.4  
176 72.2 0.0 154.0





#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.637 min Scan# 117  
Delta R.T. -0.001 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01



Tgt Ion:117 Resp: 462992  
Ion Ratio Lower Upper  
117 100  
82 61.2 44.7 67.1  
119 32.0 26.2 39.4


Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt




12/01/2021  
Supervised By :Mahesh  
Dadoda



12/01/2021




#64  
Tetrachloroethene  
Concen: 17.084 ug/l  
RT: 10.879 min Scan# 1523  
Delta R.T. 0.006 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Tgt Ion:164 Resp: 50215  
Ion Ratio Lower Upper  
164 100  
166 113.5 102.6 153.8  
129 83.5 75.4 113.0  
131 81.9 70.6 106.0







#65  
Chlorobenzene  
Concen: 17.195 ug/l  
RT: 11.667 min Scan# 1657  
Delta R.T. -0.000 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01

Tgt Ion:112 Resp: 147712  
Ion Ratio Lower Upper  
112 100  
114 33.4 26.0 39.0



Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt






12/01/2021  
Supervised By :Mahesh  
Dadoda


12/01/2021



#66  
1,1,1,2-Tetrachloroethane  
Concen: 17.513 ug/l  
RT: 11.737 min Scan# 1669  
Delta R.T. -0.001 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Tgt Ion:131 Resp: 54930  
Ion Ratio Lower Upper  
131 100  
133 95.1 48.0 144.0  
119 63.7 33.5 100.4







#67  
Ethyl Benzene  
Concen: 16.727 ug/l  
RT: 11.737 min Scan# 1669  
Delta R.T. -0.001 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01

Tgt Ion: 91 Resp: 270507  
Ion Ratio Lower Upper  
91 100  
106 30.2 24.6 36.8



Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt






12/01/2021  
Supervised By :Mahesh  
Dadoda


12/01/2021



#68  
m/p-Xylenes  
Concen: 33.616 ug/l  
RT: 11.849 min Scan# 1688  
Delta R.T. -0.001 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Tgt Ion:106 Resp: 205239  
Ion Ratio Lower Upper  
106 100  
91 205.5 160.1 240.1







#69  
o-Xylene  
Concen: 16.554 ug/l  
RT: 12.173 min Scan# 1743  
Delta R.T. -0.000 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01

Tgt Ion:106 Resp: 94992  
Ion Ratio Lower Upper  
106 100  
91 219.0 106.3 319.0



Manual Integrations  
**APPROVED**

Reviewed By :Semsettin  
Yesilyurt






12/01/2021  
Supervised By :Mahesh  
Dadoda


12/01/2021



#70  
Styrene  
Concen: 16.598 ug/l  
RT: 12.190 min Scan# 1746  
Delta R.T. -0.001 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Tgt Ion:104 Resp: 160520  
Ion Ratio Lower Upper  
104 100  
78 52.7 39.4 59.0  
103 56.5 43.4 65.0







#71  
Bromoform  
Concen: 16.858 ug/l  
RT: 12.349 min Scan# 1773  
Delta R.T. -0.006 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBSD01

Tgt Ion:173 Resp: 2739  
Ion Ratio Lower Upper  
173 100  
175 44.7 24.1 72.2  
254 0.0 0.1 0.1



Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt






12/01/2021  
Supervised By :Mahesh  
Dadoda


12/01/2021



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 13.567 min Scan# 1980  
Delta R.T. -0.000 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50


Tgt Ion:152 Resp: 211870  
Ion Ratio Lower Upper  
152 100  
115 61.9 29.8 89.3  
150 164.9 0.0 349.2





#73  
Isopropylbenzene  
Concen: 17.031 ug/l  
RT: 12.473 min Scan# 1794  
Delta R.T. -0.000 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01



Tgt Ion:105 Resp: 26128  
Ion Ratio Lower Upper  
105 100  
120 25.3 13.2 39.5


Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt




12/01/2021  
Supervised By :Mahesh  
Dadoda



12/01/2021




#74  
N-ethyl acetate  
Concen: 18.252 ug/l  
RT: 12.284 min Scan# 1762  
Delta R.T. 0.005 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Tgt Ion: 43 Resp: 59500  
Ion Ratio Lower Upper  
43 100  
70 38.1 36.6 55.0  
55 22.8 22.5 33.7  
61 21.4 20.5 30.7







#75  
1,1,2,2-Tetrachloroethane  
Concen: 17.814 ug/l  
RT: 12.726 min Scan# 1836  
Delta R.T. 0.006 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01

Tgt Ion: 83 Resp: 45004  
Ion Ratio Lower Upper  
83 100  
131 9.7 4.9 14.7  
85 62.5 32.0 96.2



Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt






12/01/2021  
Supervised By :Mahesh  
Dadoda


12/01/2021



#76  
1,2,3-Trichloropropane  
Concen: 15.353 ug/l m  
RT: 12.773 min Scan# 1845  
Delta R.T. -0.000 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50


Tgt Ion: 75 Resp: 30991  
Ion Ratio Lower Upper  
75 100  
77 350.5 158.4 475.3





#77  
Bromobenzene  
Concen: 16.819 ug/l  
RT: 12.755 min Scan# 1842  
Delta R.T. -0.000 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01



Tgt Ion:156 Resp: 55889  
Ion Ratio Lower Upper  
156 100  
77 194.3 96.6 289.8  
158 99.5 49.1 147.4


Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt




12/01/2021  
Supervised By :Mahesh  
Dadoda



12/01/2021




#78  
n-propylbenzene  
Concen: 16.704 ug/l  
RT: 12.814 min Scan# 1852  
Delta R.T. -0.000 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Tgt Ion: 91 Resp: 319119  
Ion Ratio Lower Upper  
91 100  
120 23.6 10.9 32.9







#79  
2-Chlorotoluene  
Concen: 17.072 ug/l  
RT: 12.896 min Scan# 1866  
Delta R.T. -0.006 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01

Tgt Ion: 91 Resp: 18306  
Ion Ratio Lower Upper  
91 100  
126 32.3 16.6 49.7



Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt






12/01/2021  
Supervised By :Mahesh  
Dadoda


12/01/2021



#80  
1,3,5-Trimethylbenzene  
Concen: 17.051 ug/l  
RT: 12.949 min Scan# 1875  
Delta R.T. 0.000 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Tgt Ion:105 Resp: 215710  
Ion Ratio Lower Upper  
105 100  
120 47.3 24.6 73.8







#81  
trans-1,4-Dichloro-2-butene  
Concen: 18.380 ug/l  
RT: 12.520 min Scan# 1802  
Delta R.T. -0.000 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01

Tgt Ion: 75 Resp: 1230  
Ion Ratio Lower Upper  
75 100  
53 107.8 62.6 94.0  
89 46.4 34.5 51.7



Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt






12/01/2021  
Supervised By :Mahesh  
Dadoda


12/01/2021



#82  
4-Chlorotoluene  
Concen: 16.671 ug/l  
RT: 12.996 min Scan# 1883  
Delta R.T. -0.001 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50


Tgt Ion: 91 Resp: 185581  
Ion Ratio Lower Upper  
91 100  
126 32.6 16.4 49.1





#83  
tert-Butylbenzene  
Concen: 16.735 ug/l  
RT: 13.214 min Scan# 1920  
Delta R.T. -0.000 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50


Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01



Tgt Ion:119 Resp: 181809  
Ion Ratio Lower Upper  
119 100  
91 70.0 32.5 97.5  
134 26.8 13.2 39.6


Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt




12/01/2021  
Supervised By :Mahesh  
Dadoda



12/01/2021




#84  
1,2,4-Trimethylbenzene  
Concen: 17.133 ug/l  
RT: 13.261 min Scan# 1928  
Delta R.T. -0.000 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Tgt Ion:105 Resp: 211608  
Ion Ratio Lower Upper  
105 100  
120 45.2 22.9 68.7







#85  
 sec-Butylbenzene  
 Concen: 16.855 ug/l  
 RT: 13.396 min Scan# 1950  
 Delta R.T. 0.005 min  
 Lab File: VD071083.D  
 Acq: 30 Nov 2021 12:50

Instrument :  
 MSVOA\_D  
 ClientSampleId :  
 VD1130SBS01

Tgt Ion:105 Resp: 279689  
 Ion Ratio Lower Upper  
 105 100  
 134 19.7 9.7 28.9



Manual Integrations  
 APPROVED

Reviewed By :Semsettin  
 Yesilyurt






12/01/2021  
 Supervised By :Mahesh  
 Dadoda


12/01/2021



#86  
 p-Isopropyltoluene  
 Concen: 16.829 ug/l  
 RT: 13.508 min Scan# 1970  
 Delta R.T. -0.000 min  
 Lab File: VD071083.D  
 Acq: 30 Nov 2021 12:50


Tgt Ion:119 Resp: 228399  
 Ion Ratio Lower Upper  
 119 100  
 134 25.8 13.0 38.9  
 91 25.0 11.4 34.2





#87  
1,3-Dichlorobenzene  
Concen: 17.055 ug/l  
RT: 13.508 min Scan# 1970  
Delta R.T. -0.000 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01



Tgt Ion:146 Resp: 112815  
Ion Ratio Lower Upper  
146 100  
111 43.1 19.8 59.4  
148 64.4 32.1 96.3


Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt




12/01/2021  
Supervised By :Mahesh  
Dadoda



12/01/2021




#88  
1,4-Dichlorobenzene  
Concen: 17.228 ug/l  
RT: 13.584 min Scan# 1983  
Delta R.T. -0.001 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50




Tgt Ion:146 Resp: 112906  
Ion Ratio Lower Upper  
146 100  
111 41.8 19.6 58.8  
148 62.5 31.5 94.5





#89  
n-Butylbenzene  
Concen: 16.936 ug/l  
RT: 13.831 min Scan# 2025  
Delta R.T. -0.001 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50


Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01



Tgt Ion: 91 Resp: 224370  
Ion Ratio Lower Upper  
91 100  
92 53.5 27.3 81.9  
134 24.4 12.3 36.9


Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt






12/01/2021  
Supervised By :Mahesh  
Dadoda


12/01/2021



#90  
Hexachloroethane  
Concen: 16.962 ug/l  
RT: 14.102 min Scan# 2071  
Delta R.T. 0.006 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50


Tgt Ion:117 Resp: 43639  
Ion Ratio Lower Upper  
117 100  
201 65.8 36.0 108.1





#91  
1,2-Dichlorobenzene  
Concen: 17.168 ug/l  
RT: 13.878 min Scan# 2033  
Delta R.T. -0.001 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01



Tgt Ion:146 Resp: 96574  
Ion Ratio Lower Upper  
146 100  
111 42.6 20.8 62.2  
148 62.6 31.7 95.1


Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt




12/01/2021  
Supervised By :Mahesh  
Dadoda



12/01/2021




#92  
1,2-Dibromo-3-Chloropropane  
Concen: 17.160 ug/l  
RT: 14.496 min Scan# 2138  
Delta R.T. 0.005 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50



Tgt Ion: 75 Resp: 7060  
Ion Ratio Lower Upper  
75 100  
155 79.2 40.0 120.0  
157 96.1 49.8 149.3







#93  
1,2,4-Trichlorobenzene  
Concen: 17.067 ug/l  
RT: 15.149 min Scan# 211  
Delta R.T. 0.006 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBSD01

Tgt Ion:180 Resp: 59884  
Ion Ratio Lower Upper  
180 100  
182 93.9 48.2 144.6  
145 31.7 15.4 46.2



Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt






12/01/2021  
Supervised By :Mahesh  
Dadoda


12/01/2021



#94  
Hexachlorobutadiene  
Concen: 17.848 ug/l  
RT: 15.249 min Scan# 2266  
Delta R.T. -0.006 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Tgt Ion:225 Resp: 35434  
Ion Ratio Lower Upper  
225 100  
223 60.5 31.8 95.3  
227 65.6 31.6 94.8







#95  
Naphthalene  
Concen: 16.366 ug/l  
RT: 15.384 min Scan# 2113  
Delta R.T. -0.001 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Instrument :  
MSVOA\_D  
ClientSampleId :  
VD1130SBS01

Tgt Ion:128 Resp: 10123  
Ion Ratio Lower Upper  
128 100  
127 14.3 10.7 16.1  
129 11.0 8.7 13.1



Manual Integrations  
APPROVED

Reviewed By :Semsettin  
Yesilyurt



12/01/2021  
Supervised By :Mahesh  
Dadoda

12/01/2021



#96  
1,2,3-Trichlorobenzene  
Concen: 16.746 ug/l  
RT: 15.578 min Scan# 2322  
Delta R.T. -0.001 min  
Lab File: VD071083.D  
Acq: 30 Nov 2021 12:50

Tgt Ion:180 Resp: 49907  
Ion Ratio Lower Upper  
180 100  
182 98.6 46.7 140.1  
145 33.4 15.9 47.6

