

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120221\
 Data File : VD071116.D
 Acq On : 02 Dec 2021 14:20
 Operator : VA/SY
 Sample : M4876-01
 Misc : 5.09G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CL-02-120121

Quant Time: Dec 03 00:37:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	2535	50.000	ug/l	#-0.01
34) 1,4-Difluorobenzene	8.861	114	4618	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	3685	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	1661	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	1710	57.488	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 114.980%	
35) Dibromofluoromethane	7.914	113	819	27.851	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 55.700%	
50) Toluene-d8	10.338	98	4704	42.034	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 84.060%	
62) 4-Bromofluorobenzene	12.626	95	1405	36.219	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 72.440%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120221\
Data File : VD071116.D
Acq On : 02 Dec 2021 14:20
Operator : VA/SY
Sample : M4876-01
Misc : 5.09G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CL-02-120121

Quant Time: Dec 03 00:37:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

