

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD120318\
 Data File : VD060467.D
 Acq On : 3 Dec 2018 11:04
 Operator : VA/AP
 Sample : VD1203SBL01
 Misc : 5.00g/5ml/MSVOA D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1203SBL01

Quant Time: Dec 04 02:29:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D110618S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 19 16:31:59 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.43	168	1376998	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	7.47	114	1885247	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.30	117	1499230	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.38	152	739817	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.82	65	512924	52.02	ug/l	0.02
Spiked Amount 50.000			Recovery	=	104.04%	
35) Dibromofluoromethane	6.35	113	752906	53.56	ug/l	0.02
Spiked Amount 50.000			Recovery	=	107.12%	
50) Toluene-d8	9.48	98	2137898	51.96	ug/l	0.02
Spiked Amount 50.000			Recovery	=	103.92%	
62) 4-Bromofluorobenzene	12.44	95	778810	52.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.48%	
Target Compounds						
20) Methylene Chloride	3.76	84	34564	2.278	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD120318\
Data File : VD060467.D
Acq On : 3 Dec 2018 11:04
Operator : VA/AP
Sample : VD1203SBL01
Misc : 5.00g/5ml/MSVOA D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1203SBL01

Quant Time: Dec 04 02:29:58 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D110618S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 19 16:31:59 2018
Response via : Initial Calibration

