

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD120320\
 Data File : VD067861.D
 Acq On : 03 Dec 2020 11:55
 Operator : VA/SY
 Sample : L4909-02
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ROCK-HILL-SLOPE-EMBANKMENT-

Quant Time: Dec 03 14:25:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D120120S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 02 08:05:58 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	227759	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	322293	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	280704	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	133286	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.34	65	107783	44.08	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	88.16%
35) Dibromofluoromethane	7.92	113	95477	44.27	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	88.54%
50) Toluene-d8	10.34	98	365453	47.07	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	94.14%
62) 4-Bromofluorobenzene	12.64	95	113610	42.33	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	84.66%

Target Compounds

					Qvalue
20) Methylene Chloride	4.97	84	25971	9.983	ug/l 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD120320\
Data File : VD067861.D
Acq On : 03 Dec 2020 11:55
Operator : VA/SY
Sample : L4909-02
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
ROCK-HILL-SLOPE-EMBANKMENT-

Quant Time: Dec 03 14:25:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D120120S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 02 08:05:58 2020
Response via : Initial Calibration

