

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD120320\
 Data File : VD067873.D
 Acq On : 03 Dec 2020 17:54
 Operator : VA/SY
 Sample : L4935-01RE
 Misc : 7.33G/5.00ml/MSVOA D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OK-02-12220RE

Quant Time: Dec 04 01:55:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D120120S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 02 08:05:58 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	104613	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	145221	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	137792	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	67081	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.34	65	65676	58.48	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.96%
35) Dibromofluoromethane	7.92	113	50045	51.50	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.00%
50) Toluene-d8	10.34	98	166748	47.67	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.34%
62) 4-Bromofluorobenzene	12.63	95	55223	45.66	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	91.32%

Target Compounds					Qvalue
95) Naphthalene	15.40	128	16529	7.568 ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD120320\
Data File : VD067873.D
Acq On : 03 Dec 2020 17:54
Operator : VA/SY
Sample : L4935-01RE
Misc : 7.33G/5.00ml/MSVOA D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OK-02-12220RE

Quant Time: Dec 04 01:55:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D120120S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 02 08:05:58 2020
Response via : Initial Calibration

