

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120321\
 Data File : VD071135.D
 Acq On : 03 Dec 2021 15:21
 Operator : VA/SY
 Sample : M4910-19RE
 Misc : 5.40G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SP-1-4

Quant Time: Dec 04 03:50:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	294252	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	531745	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	561279	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	262791	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	173214	50.168	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 100.340%	
35) Dibromofluoromethane	7.908	113	168543	49.775	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 99.560%	
50) Toluene-d8	10.332	98	681549	52.891	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 105.780%	
62) 4-Bromofluorobenzene	12.626	95	268814	60.182	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 120.360%	
Target Compounds						
16) Acetone	4.144	43	20019	49.942	ug/l	Qvalue # 72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120321\
Data File : VD071135.D
Acq On : 03 Dec 2021 15:21
Operator : VA/SY
Sample : M4910-19RE
Misc : 5.40G/5.00ml/MSVOA_D/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SP-1-4

Quant Time: Dec 04 03:50:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

