

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120321\
 Data File : VD071136.D
 Acq On : 03 Dec 2021 15:49
 Operator : VA/SY
 Sample : M4920-01RE
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 FDR-20211202-WC-SRE

Quant Time: Dec 04 03:50:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	3941	50.000	ug/l	#-0.01
34) 1,4-Difluorobenzene	8.867	114	6912	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	7226	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	2874	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	3378	73.049	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 146.100%	
35) Dibromofluoromethane	7.914	113	1685	38.282	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 76.560%	
50) Toluene-d8	10.332	98	8615	51.433	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 102.860%	
62) 4-Bromofluorobenzene	12.626	95	3424	58.972	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 117.940%	
Target Compounds						
83) tert-Butylbenzene	13.214	119	3916	26.572	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120321\
Data File : VD071136.D
Acq On : 03 Dec 2021 15:49
Operator : VA/SY
Sample : M4920-01RE
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
FDR-20211202-WC-SRE

Quant Time: Dec 04 03:50:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

