

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD120418\
 Data File : VD060486.D
 Acq On : 4 Dec 2018 14:39
 Operator : VA/AP
 Sample : J6201-01
 Misc : 6.07g/5ml/MSVOA D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 72-11932-72-11967

Quant Time: Dec 05 02:48:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D110618S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 19 16:31:59 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.41	168	1377788	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	7.45	114	1891684	50.00	ug/l	-0.01
63) Chlorobenzene-d5	11.29	117	1423796	50.00	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	13.37	152	565811	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.79	65	496139	50.29	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	100.58%	
35) Dibromofluoromethane	6.33	113	696652	49.39	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	98.78%	
50) Toluene-d8	9.45	98	1774024	42.97	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	85.94%	
62) 4-Bromofluorobenzene	12.43	95	569777	38.09	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	76.18%	
Target Compounds						
16) Acetone	3.17	43	65854	38.484	ug/l	95
20) Methylene Chloride	3.73	84	58654	3.864	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD120418\
Data File : VD060486.D
Acq On : 4 Dec 2018 14:39
Operator : VA/AP
Sample : J6201-01
Misc : 6.07g/5ml/MSVOA D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
72-11932-72-11967

Quant Time: Dec 05 02:48:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D110618S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 19 16:31:59 2018
Response via : Initial Calibration

