

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD120418\
 Data File : VD060487.D
 Acq On : 4 Dec 2018 15:07
 Operator : VA/AP
 Sample : J6198-01
 Misc : 8.42g/5ml/MSVOA D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SA-06-120318

Quant Time: Dec 05 02:52:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D110618S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 19 16:31:59 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.42	168	170999	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	7.46	114	274424	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.28	117	256568	50.00	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	13.38	152	130761	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.80	65	116465	95.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	190.24%	
35) Dibromofluoromethane	6.33	113	116803	57.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.16%	
50) Toluene-d8	9.46	98	238133	39.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.52%	
62) 4-Bromofluorobenzene	12.43	95	108444	49.97	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	99.94%	
Target Compounds						
16) Acetone	3.18	43	15288	71.984	ug/l #	85
20) Methylene Chloride	3.75	84	26672	14.157	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD120418\
Data File : VD060487.D
Acq On : 4 Dec 2018 15:07
Operator : VA/AP
Sample : J6198-01
Misc : 8.42g/5ml/MSVOA D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SA-06-120318

Quant Time: Dec 05 02:52:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D110618S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 19 16:31:59 2018
Response via : Initial Calibration

