

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD120420\
 Data File : VD067883.D
 Acq On : 04 Dec 2020 14:10
 Operator : VA/SY
 Sample : L4599-01
 Misc : 4.94G/5.00ml/MSVOA D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TT-034-IDWSO-20201203

Quant Time: Dec 05 00:23:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D120120S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 02 08:05:58 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	209338	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	285746	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	235466	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	99552	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.34	65	88884	39.55	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	79.10%
35) Dibromofluoromethane	7.91	113	82905	43.36	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	86.72%
50) Toluene-d8	10.34	98	317947	46.19	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	92.38%
62) 4-Bromofluorobenzene	12.64	95	88813	37.32	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	74.64%

Target Compounds					Qvalue
5) Bromomethane	2.78	94	1920	1.624 ug/l #	58

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD120420\
Data File : VD067883.D
Acq On : 04 Dec 2020 14:10
Operator : VA/SY
Sample : L4599-01
Misc : 4.94G/5.00ml/MSVOA D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TT-034-IDWSO-20201203

Quant Time: Dec 05 00:23:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D120120S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 02 08:05:58 2020
Response via : Initial Calibration

