

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD120519\
 Data File : VD064434.D
 Acq On : 05 Dec 2019 14:10
 Operator : VA/SY
 Sample : K6147-07
 Misc : 5.53G/5.00ml/MSVOA D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P001-WC007

Quant Time: Dec 06 08:13:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D112719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 29 07:18:03 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	49096	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	83088	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	83278	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	42627	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	34054	76.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	153.26%	
35) Dibromofluoromethane	7.93	113	28786	54.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.98%	
50) Toluene-d8	10.34	98	100214	50.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.64%	
62) 4-Bromofluorobenzene	12.64	95	36267	57.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.46%	
Target Compounds						
16) Acetone	4.17	43	6812	75.855	ug/l	Qvalue 100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD120519\
Data File : VD064434.D
Acq On : 05 Dec 2019 14:10
Operator : VA/SY
Sample : K6147-07
Misc : 5.53G/5.00ml/MSVOA D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleID :
P001-WC007

Quant Time: Dec 06 08:13:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D112719S.M
Quant Title : SW846 8260
QLast Update : Fri Nov 29 07:18:03 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.98 min Scan# 1031

Delta R.T. 0.01 min

Lab File: VD064434.D

Acq: 05 Dec 2019 14:10

Instrument :

MSVOA_D

ClientSampleId :


P001-WC007

Tot Ion:168 Resp: 49096

Ion Ratio Lower Upper




168 100

99 51.1 40.7 61.1



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VDC



#16

Acetone

Concen: 75.855 ug/l

RT: 4.17 min Scan# 382

Delta R.T. -0.01 min

Lab File: VD064434.D


Acq: 05 Dec 2019 14:10

Tgt Ion: 43 Resp: 6812

Ion Ratio Lower Upper



43 100


58 28.4 22.6 33.8



Abundance Ion 43.00 (42.70 to 43.70): VDC

Ion 58.00 (57.70 to 58.70): VDC









#33
 1,2-Dichloroethane-d4
 Concen: 76.631 ug/l
 RT: 8.33 min Scan# 1090
 Delta R.T. -0.01 min
 Lab File: VD064434.D
 Acq: 05 Dec 2019 14:10




Instrument :
 MSVOA_D
 ClientSampleId :
 P001-WC007


Tot Ion: 65 Resp: 34054
 Ion Ratio Lower Upper
 65 100
 67 52.9 0.0 102.0



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.87 min Scan# 1181
 Delta R.T. -0.00 min
 Lab File: VD064434.D
 Acq: 05 Dec 2019 14:10

Tgt Ion: 114 Resp: 83088
 Ion Ratio Lower Upper
 114 100
 63 18.5 0.0 40.0
 88 14.7 0.0 29.2









#35
 Dibromofluoromethane
 Concen: 54.988 ug/l
 RT: 7.93 min Scan# 1021
 Delta R.T. 0.01 min
 Lab File: VD064434.D
 Acq: 05 Dec 2019 14:10




Instrument :
 MSVOA_D
 ClientSampleId :
 P001-WC007


Tot Ion:113 Resp: 28786
 Ion Ratio Lower Upper
 113 100
 111 102.6 79.4 119.2
 192 23.9 18.6 28.0



#50
 Toluene-d8
 Concen: 50.820 ug/l
 RT: 10.34 min Scan# 1431
 Delta R.T. -0.01 min
 Lab File: VD064434.D
 Acq: 05 Dec 2019 14:10

Tgt Ion: 98 Resp: 100214
 Ion Ratio Lower Upper
 98 100
 100 63.5 52.6 79.0









#62
4-Bromofluorobenzene
Concen: 57.228 ug/l
RT: 12.64 min Scan# 1822
Delta R.T. -0.00 min
Lab File: VD064434.D
Acq: 05 Dec 2019 14:10




Instrument :
MSVOA_D
ClientSampleId :
P001-WC007


Tot Ion: 95 Resp: 36267
Ion Ratio Lower Upper
95 100
174 98.2 0.0 197.2
176 91.1 0.0 189.6



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.65 min Scan# 1654
Delta R.T. -0.00 min
Lab File: VD064434.D
Acq: 05 Dec 2019 14:10

Tgt Ion: 117 Resp: 83278
Ion Ratio Lower Upper
117 100
82 50.9 41.4 62.2
119 33.7 26.2 39.2





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.58 min Scan# 1982

Delta R.T. -0.01 min

Lab File: VD064434.D

Acq: 05 Dec 2019 14:10

Instrument :

MSVOA_D

ClientSampleId :

P001-WC007


Tot Ion:152 Resp: 42627

Ion Ratio Lower Upper

152 100

115 53.0 26.7 80.1

150 157.0 0.0 347.2



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

