

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD120618\
 Data File : VD060501.D
 Acq On : 6 Dec 2018 13:15
 Operator : VA/AP
 Sample : J6195-01
 Misc : 6.6g/5ml/MSVOA D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-01-120518-A

Quant Time: Dec 07 03:54:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D120518S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 06 01:49:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.41	168	1537949	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	7.44	114	2112163	50.00	ug/l	-0.01
63) Chlorobenzene-d5	11.28	117	1676226	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.37	152	731954	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.78	65	542121	49.73	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	99.46%	
35) Dibromofluoromethane	6.33	113	654733	42.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.40%	
50) Toluene-d8	9.45	98	1711134	38.11	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	76.22%	
62) 4-Bromofluorobenzene	12.42	95	593697	36.96	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	73.92%	
Target Compounds						
					Qvalue	
16) Acetone	3.17	43	67442	38.597	ug/l	94
20) Methylene Chloride	3.73	84	182499	7.580	ug/l	98
25) 2-Butanone	5.62	43	12930	3.761	ug/l #	83
52) Toluene	9.55	92	712499	21.604	ug/l	96
67) Ethyl Benzene	11.43	91	143588	2.374	ug/l	98
68) m/p-Xylenes	11.56	106	308254	13.775	ug/l	100
69) o-Xylene	11.93	106	124640	5.845	ug/l	92
95) Naphthalene	14.97	128	36414	1.944	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD120618\
Data File : VD060501.D
Acq On : 6 Dec 2018 13:15
Operator : VA/AP
Sample : J6195-01
Misc : 6.6q/5ml/MSVOA D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-01-120518-A

Quant Time: Dec 07 03:54:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D120518S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 06 01:49:20 2018
Response via : Initial Calibration

