

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120721\
 Data File : VD071182.D
 Acq On : 07 Dec 2021 14:40
 Operator : VA/SY
 Sample : M4959-04
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-1D

Quant Time: Dec 08 02:49:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	15393	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	28461	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	33361	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	15867	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	13994	77.478	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	154.960%
35) Dibromofluoromethane	7.920	113	7982	44.042	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	88.080%
50) Toluene-d8	10.337	98	33849	49.078	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.160%
62) 4-Bromofluorobenzene	12.625	95	14919	62.403	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	124.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120721\
Data File : VD071182.D
Acq On : 07 Dec 2021 14:40
Operator : VA/SY
Sample : M4959-04
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-1D

Quant Time: Dec 08 02:49:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

