

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120721\
 Data File : VD071187.D
 Acq On : 07 Dec 2021 17:00
 Operator : VA/SY
 Sample : M4959-16
 Misc : 5.03G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-5D

Quant Time: Dec 08 02:51:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	153542	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.861	114	273304	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	316623	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	158760	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	111991	62.161	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 124.320%	
35) Dibromofluoromethane	7.914	113	93736	53.860	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 107.720%	
50) Toluene-d8	10.338	98	357343	53.955	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 107.900%	
62) 4-Bromofluorobenzene	12.626	95	157539	68.621	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 137.240%	
Target Compounds						
						Qvalue
16) Acetone	4.162	43	5664	41.177	ug/l	95
20) Methylene Chloride	4.973	84	11005	2.048	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120721\
Data File : VD071187.D
Acq On : 07 Dec 2021 17:00
Operator : VA/SY
Sample : M4959-16
Misc : 5.03G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-5D

Quant Time: Dec 08 02:51:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

