

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD120820\
 Data File : VD067905.D
 Acq On : 08 Dec 2020 10:24
 Operator : VA/SY
 Sample : VD1208SBL01
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1208SBL01

Quant Time: Dec 09 04:33:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D120120S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 02 08:05:58 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.99	168	248751	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	353418	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	305706	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	149511	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.33	65	114561	42.90	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	85.80%
35) Dibromofluoromethane	7.92	113	100325	42.42	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	84.84%
50) Toluene-d8	10.34	98	387589	45.53	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	91.06%
62) 4-Bromofluorobenzene	12.64	95	127927	43.46	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	86.92%

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD120820\
Data File : VD067905.D
Acq On : 08 Dec 2020 10:24
Operator : VA/SY
Sample : VD1208SBL01
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1208SBL01

Quant Time: Dec 09 04:33:00 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D120120S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 02 08:05:58 2020
Response via : Initial Calibration

