

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120821\
 Data File : VD071200.D
 Acq On : 08 Dec 2021 12:44
 Operator : VA/SY
 Sample : M4949-01
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VNJ-221

Quant Time: Dec 09 10:14:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.984	168	12262	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	20062	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	22739	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	11197	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	10360	72.004	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	144.000%
35) Dibromofluoromethane	7.920	113	6166	48.265	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.540%
50) Toluene-d8	10.337	98	24982	51.386	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.780%
62) 4-Bromofluorobenzene	12.625	95	10884	64.585	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	129.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120821\
Data File : VD071200.D
Acq On : 08 Dec 2021 12:44
Operator : VA/SY
Sample : M4949-01
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VNJ-221

Quant Time: Dec 09 10:14:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

