

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120821\
 Data File : VD071205.D
 Acq On : 08 Dec 2021 15:03
 Operator : VA/SY
 Sample : M4959-10RE
 Misc : 4.99G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-3SRE

Quant Time: Dec 09 10:16:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	62459	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	106806	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	125487	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	63574	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	48388	66.024	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 132.040%	
35) Dibromofluoromethane	7.914	113	37416	55.013	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 110.020%	
50) Toluene-d8	10.338	98	143187	55.322	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 110.640%	
62) 4-Bromofluorobenzene	12.626	95	60260	67.166	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 134.340%	
Target Compounds						Qvalue
16) Acetone	4.168	43	7381	64.052	ug/l #	83
20) Methylene Chloride	4.968	84	4473	2.042	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120821\
Data File : VD071205.D
Acq On : 08 Dec 2021 15:03
Operator : VA/SY
Sample : M4959-10RE
Misc : 4.99G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-3SRE

Quant Time: Dec 09 10:16:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

