

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120821\
 Data File : VD071207.D
 Acq On : 08 Dec 2021 15:59
 Operator : VA/SY
 Sample : M4959-19RE
 Misc : 5.03G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-3DRE

Quant Time: Dec 09 10:17:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	25401	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	43577	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	50644	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	25865	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	22096	74.135	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	148.260%
35) Dibromofluoromethane	7.908	113	15854	57.133	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.260%
50) Toluene-d8	10.332	98	56699	53.692	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.380%
62) 4-Bromofluorobenzene	12.620	95	25741	70.321	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	140.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120821\
Data File : VD071207.D
Acq On : 08 Dec 2021 15:59
Operator : VA/SY
Sample : M4959-19RE
Misc : 5.03G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-3DRE

Quant Time: Dec 09 10:17:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

