

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120821\
 Data File : VD071210.D
 Acq On : 08 Dec 2021 17:22
 Operator : VA/SY
 Sample : M4966-19
 Misc : 6.25G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 P14-11

Quant Time: Dec 09 10:18:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	83626	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	144861	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	172725	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	88090	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	67043	68.324	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 136.640%	
35) Dibromofluoromethane	7.914	113	52404	56.809	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 113.620%	
50) Toluene-d8	10.338	98	189935	54.106	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 108.220%	
62) 4-Bromofluorobenzene	12.626	95	85553	70.307	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 140.620%	
Target Compounds						Qvalue
20) Methylene Chloride	4.967	84	5422	1.418	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120821\
Data File : VD071210.D
Acq On : 08 Dec 2021 17:22
Operator : VA/SY
Sample : M4966-19
Misc : 6.25G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
P14-11

Quant Time: Dec 09 10:18:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

