

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120821\
 Data File : VD071212.D
 Acq On : 08 Dec 2021 18:17
 Operator : VA/SY
 Sample : M4964-01
 Misc : 6.91G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 LAUNDRYROOM-PIT-BOTTOM

Quant Time: Dec 09 10:19:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	122841	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	216885	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	255778	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	127761	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	86546	60.043	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	120.080%
35) Dibromofluoromethane	7.914	113	73623	53.307	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.620%
50) Toluene-d8	10.338	98	287450	54.692	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.380%
62) 4-Bromofluorobenzene	12.626	95	126113	69.222	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	138.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120821\
Data File : VD071212.D
Acq On : 08 Dec 2021 18:17
Operator : VA/SY
Sample : M4964-01
Misc : 6.91G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
LAUNDRYROOM-PIT-BOTTOM

Quant Time: Dec 09 10:19:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

