

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120821\
 Data File : VD071201.D
 Acq On : 08 Dec 2021 13:12
 Operator : VA/SY
 Sample : M4955-01
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 2022

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Title : SW846 8260

Signal : TIC: VD071201.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	2.121	31	34	35	rBV	1659	1619	2.41%	0.308%
2	2.221	45	51	52	rBV2	3684	5019	7.48%	0.954%
3	2.538	103	105	109	rBV4	2887	4607	6.86%	0.876%
4	2.621	116	119	121	rBV2	4055	4452	6.63%	0.846%
5	2.732	136	138	139	rBV2	3290	1895	2.82%	0.360%
6	2.785	145	147	150	rBV4	2614	3427	5.11%	0.651%
7	2.838	154	156	158	rVB2	2265	1624	2.42%	0.309%
8	2.979	179	180	183	rBV2	1269	1218	1.81%	0.232%
9	3.044	189	191	193	rVB	1833	1664	2.48%	0.316%
10	3.068	193	195	196	rBV	1677	1267	1.89%	0.241%
11	3.150	208	209	211	rBV	1951	1222	1.82%	0.232%
12	3.232	219	223	224	rVB	1431	1726	2.57%	0.328%
13	3.391	249	250	253	rVV	1323	1379	2.05%	0.262%
14	3.426	255	256	259	rVB2	1404	1290	1.92%	0.245%
15	3.450	259	260	264	rBV2	1213	1522	2.27%	0.289%
16	3.662	292	296	298	rBV2	1425	1458	2.17%	0.277%
17	3.873	329	332	335	rBV2	828	1414	2.11%	0.269%
18	3.979	348	350	353	rBV2	1217	1835	2.73%	0.349%
19	4.097	367	370	372	rBV2	1266	1240	1.85%	0.236%
20	4.150	376	379	380	rBV2	1369	1309	1.95%	0.249%
21	4.221	389	391	396	rVB2	1561	1735	2.58%	0.330%
22	4.344	409	412	413	rBV2	1326	1251	1.86%	0.238%
23	4.909	507	508	512	rVB2	1151	1315	1.96%	0.250%
24	4.938	512	513	516	rBV2	1591	1677	2.50%	0.319%
25	5.020	526	527	532	rBV	1367	1308	1.95%	0.249%
26	5.126	541	545	548	rBV2	1545	2333	3.48%	0.443%
27	5.550	614	617	620	rBV2	1756	2080	3.10%	0.395%
28	5.926	678	681	682	rVB2	1371	1378	2.05%	0.262%
29	6.638	797	802	803	rBV2	1027	1616	2.41%	0.307%
30	6.703	811	813	814	rVB	1871	1321	1.97%	0.251%
31	6.750	819	821	823	rVV2	1338	1398	2.08%	0.266%
32	6.809	826	831	832	rVV2	1216	1876	2.79%	0.357%
33	7.109	879	882	883	rBV2	1375	1252	1.87%	0.238%
34	7.156	885	890	891	rBV	1873	2655	3.96%	0.505%
35	7.185	892	895	896	rVV2	2068	1732	2.58%	0.329%
36	7.385	925	929	931	rBV2	1166	1972	2.94%	0.375%

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120821\
 Data File : VD071201.D
 Acq On : 08 Dec 2021 13:12
 Operator : VA/SY
 Sample : M4955-01
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 2022

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Title : SW846 8260

37	7.532	951	954	956	rVB2	1858	1874	2.79%	0.356%
38	7.673	975	978	979	rBV	1347	1631	2.43%	0.310%
39	7.697	981	982	987	rVB2	1330	1249	1.86%	0.237%
40	7.861	1007	1010	1011	rBV2	1879	1765	2.63%	0.336%
41	7.914	1011	1019	1024	rVV3	8725	20516	30.57%	3.900%
42	7.973	1024	1029	1038	rVB2	15291	33762	50.30%	6.418%
43	8.138	1056	1057	1060	rVV2	1212	1228	1.83%	0.233%
44	8.332	1083	1090	1095	rVV2	12281	23550	35.09%	4.477%
45	8.438	1106	1108	1111	rVB	1407	1606	2.39%	0.305%
46	8.544	1124	1126	1129	rVB	1408	1256	1.87%	0.239%
47	8.567	1129	1130	1132	rBV	1851	1287	1.92%	0.245%
48	8.656	1142	1145	1148	rBV2	1288	1880	2.80%	0.357%
49	8.726	1153	1157	1159	rVV2	1166	1591	2.37%	0.302%
50	8.861	1173	1180	1188	rVB2	23931	44818	66.77%	8.520%
51	8.973	1197	1199	1203	rVB2	1092	1351	2.01%	0.257%
52	9.085	1215	1218	1223	rVB2	1321	2333	3.48%	0.443%
53	9.250	1244	1246	1250	rBV	1377	1259	1.88%	0.239%
54	9.538	1290	1295	1299	rVV	867	1973	2.94%	0.375%
55	9.591	1303	1304	1307	rBV2	1317	1347	2.01%	0.256%
56	9.867	1349	1351	1353	rBV2	1445	1330	1.98%	0.253%
57	9.985	1369	1371	1373	rVV	1685	1481	2.21%	0.282%
58	10.126	1394	1395	1397	rBV	1235	1250	1.86%	0.238%
59	10.185	1401	1405	1410	rVV2	1536	2343	3.49%	0.445%
60	10.238	1410	1414	1417	rVV	1343	1600	2.38%	0.304%
61	10.338	1423	1431	1438	rBV2	32642	63953	95.28%	12.157%
62	10.485	1453	1456	1457	rBV	1604	1457	2.17%	0.277%
63	10.703	1489	1493	1499	rVB2	1537	2299	3.43%	0.437%
64	10.814	1510	1512	1514	rBV2	1669	1451	2.16%	0.276%
65	10.844	1514	1517	1520	rVB2	1225	1464	2.18%	0.278%
66	11.079	1555	1557	1560	rBV2	2175	1959	2.92%	0.372%
67	11.320	1595	1598	1601	rVB2	1060	1342	2.00%	0.255%
68	11.491	1626	1627	1630	rBV	1371	1224	1.82%	0.233%
69	11.638	1646	1652	1659	rVB2	40675	67121	100.00%	12.759%
70	11.791	1675	1678	1679	rBV	1667	1301	1.94%	0.247%
71	11.967	1707	1708	1710	rVB	1939	1225	1.83%	0.233%
72	12.238	1753	1754	1757	rVB2	1609	1387	2.07%	0.264%
73	12.273	1757	1760	1761	rBV	1158	1309	1.95%	0.249%
74	12.573	1808	1811	1814	rVB2	1003	1425	2.12%	0.271%
75	12.626	1814	1820	1827	rBV3	30128	50722	75.57%	9.642%
76	12.773	1844	1845	1848	rBV2	1531	1627	2.42%	0.309%

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120821\
 Data File : VD071201.D
 Acq On : 08 Dec 2021 13:12
 Operator : VA/SY
 Sample : M4955-01
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 2022

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Title : SW846 8260

77	12.861	1859	1860	1863	rVB2	1940	1388	2.07%	0.264%
78	13.185	1913	1915	1918	rVB	2150	1551	2.31%	0.295%
79	13.444	1957	1959	1962	rVB2	1548	1380	2.06%	0.262%
80	13.567	1974	1980	1987	rVB2	38548	63866	95.15%	12.141%
81	13.726	2005	2007	2010	rBV	1466	1344	2.00%	0.255%
82	13.955	2043	2046	2048	rVB	1434	1389	2.07%	0.264%
83	13.985	2048	2051	2052	rBV2	1303	1498	2.23%	0.285%
84	14.067	2062	2065	2067	rBV2	1033	1272	1.90%	0.242%
85	14.296	2101	2104	2106	rBV2	1477	1838	2.74%	0.349%
86	14.326	2107	2109	2111	rVV3	2022	1433	2.13%	0.272%
87	14.567	2148	2150	2152	rVB2	1343	1229	1.83%	0.234%
88	14.626	2157	2160	2163	rVV3	1597	1713	2.55%	0.326%
89	15.008	2223	2225	2227	rBV	1869	1640	2.44%	0.312%
90	15.043	2229	2231	2233	rBV2	2030	1671	2.49%	0.318%
91	15.249	2265	2266	2271	rVB3	2023	1279	1.91%	0.243%
92	15.291	2271	2273	2276	rBV3	1309	1391	2.07%	0.264%
93	15.385	2286	2289	2294	rBV4	2266	3209	4.78%	0.610%
94	15.467	2301	2303	2306	rVB2	1939	1812	2.70%	0.344%
95	15.879	2371	2373	2375	rBV3	1932	1986	2.96%	0.378%
96	15.926	2380	2381	2382	rBV	2277	1266	1.89%	0.241%
97	16.061	2401	2404	2407	rVB3	2051	2452	3.65%	0.466%
98	16.502	2477	2479	2480	rBV	2157	1340	2.00%	0.255%
99	16.749	2518	2521	2522	rVV3	2862	2776	4.14%	0.528%
100	16.773	2524	2525	2527	rVB2	2536	1695	2.53%	0.322%

Sum of corrected areas: 526050

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120821\
Data File : VD071201.D
Acq On : 08 Dec 2021 13:12
Operator : VA/SY
Sample : M4955-01
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
2022

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260

TIC Library :
TIC Integration Parameters: RTEINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120821\
Data File : VD071201.D
Acq On : 08 Dec 2021 13:12
Operator : VA/SY
Sample : M4955-01
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
2022

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260

TIC Library :
TIC Integration Parameters: RTEINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc