

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120823\
 Data File : VD077793.D
 Acq On : 08 Dec 2023 15:59
 Operator : JC/SY
 Sample : 05629-01
 Misc : 5.05G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VNJ-216

Quant Time: Dec 08 16:10:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 08 13:33:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	171493	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	356485	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	325475	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	144112	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	97507	48.252	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.500%
35) Dibromofluoromethane	7.804	113	116993	46.852	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.700%
50) Toluene-d8	10.269	98	441540	44.149	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	88.300%
62) 4-Bromofluorobenzene	12.575	95	127136	42.520	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	85.040%
Target Compounds						
20) Methylene Chloride	4.810	84	15158	1.650	ug/l	Qvalue # 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120823\
Data File : VD077793.D
Acq On : 08 Dec 2023 15:59
Operator : JC/SY
Sample : 05629-01
Misc : 5.05G/5.0ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VNJ-216

Quant Time: Dec 08 16:10:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
Quant Title : SW846 8260
QLast Update : Fri Dec 08 13:33:18 2023
Response via : Initial Calibration

