

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120823\
 Data File : VD077798.D
 Acq On : 08 Dec 2023 18:31
 Operator : JC/SY
 Sample : 05671-05
 Misc : 5.04G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-35

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 12/11/2023
 Supervised By :Mahesh Dadoda 12/11/2023

Quant Time: Dec 08 23:13:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 08 13:33:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	4700	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.781	114	9952	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	9474	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	4661	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	4012	72.441	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 144.880%	
35) Dibromofluoromethane	7.811	113	4463m	64.022	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 128.040%	
50) Toluene-d8	10.275	98	12253	43.886	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 87.780%	
62) 4-Bromofluorobenzene	12.569	95	3905	46.781	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 93.560%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120823\
Data File : VD077798.D
Acq On : 08 Dec 2023 18:31
Operator : JC/SY
Sample : 05671-05
Misc : 5.04G/5.0ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-35

Quant Time: Dec 08 23:13:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
Quant Title : SW846 8260
QLast Update : Fri Dec 08 13:33:18 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 12/11/2023
Supervised By :Mahesh Dadoda 12/11/2023

