

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120921\
 Data File : VD071224.D
 Acq On : 09 Dec 2021 13:21
 Operator : VA/SY
 Sample : M4964-01RE
 Misc : 7.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 LAUNDRYROOM-PIT-BOTTOMRE

Quant Time: Dec 10 06:43:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	21135	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	8.861	114	36156	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	42281	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	21580	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	18070	72.86	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	145.72%
35) Dibromofluoromethane	7.914	113	12722	55.26	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.52%
50) Toluene-d8	10.332	98	47891	54.66	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.32%
62) 4-Bromofluorobenzene	12.626	95	20577	67.75	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	135.50%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120921\
Data File : VD071224.D
Acq On : 09 Dec 2021 13:21
Operator : VA/SY
Sample : M4964-01RE
Misc : 7.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
LAUNDRYROOM-PIT-BOTTOMRE

Quant Time: Dec 10 06:43:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

