

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120921\
 Data File : VD071235.D
 Acq On : 09 Dec 2021 18:26
 Operator : VA/SY
 Sample : M4967-01
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-7S

Quant Time: Dec 10 06:47:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	6189	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.867	114	10889	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	10952	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	5385	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	5195	71.54	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	143.08%
35) Dibromofluoromethane	7.920	113	2885	41.61	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	83.22%
50) Toluene-d8	10.344	98	12227	46.34	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	92.68%
62) 4-Bromofluorobenzene	12.626	95	5146	56.26	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	112.52%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120921\
Data File : VD071235.D
Acq On : 09 Dec 2021 18:26
Operator : VA/SY
Sample : M4967-01
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-7S

Quant Time: Dec 10 06:47:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

