

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120921\
 Data File : VD071237.D
 Acq On : 09 Dec 2021 19:21
 Operator : VA/SY
 Sample : M4967-07
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-8S

Quant Time: Dec 10 06:47:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	33466	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	56996	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	66306	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	35439	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	27277	69.46	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 138.92%	
35) Dibromofluoromethane	7.908	113	18989	52.32	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 104.64%	
50) Toluene-d8	10.337	98	72642	52.59	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 105.18%	
62) 4-Bromofluorobenzene	12.625	95	33081	69.10	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 138.20%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD120921\
Data File : VD071237.D
Acq On : 09 Dec 2021 19:21
Operator : VA/SY
Sample : M4967-07
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-8S

Quant Time: Dec 10 06:47:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

