


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD121019\
Data File : VD064481.D
Acq On : 10 Dec 2019 09:46
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D112719S.M
Title : SW846 8260
Last Update : Fri Nov 29 07:18:03 2019



AutoFind: Scans 1821, 1822, 1823; Background Corrected with Scan 1812

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.1	23389	PASS
75	95	30	60	51.7	60032	PASS
95	95	100	100	100.0	116200	PASS
96	95	5	9	6.8	7876	PASS
173	174	0.00	2	1.4	1543	PASS
174	95	50	100	95.4	110885	PASS
175	174	5	9	7.4	8184	PASS
176	174	95	101	97.2	107746	PASS
177	176	5	9	7.2	7708	PASS