

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121021\
 Data File : VD071254.D
 Acq On : 10 Dec 2021 16:23
 Operator : VA/SY
 Sample : M4990-04
 Misc : 6.80G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 R3-7

Quant Time: Dec 11 04:40:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	267180	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	461948	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	501876	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	245485	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	144054	45.95	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.90%
35) Dibromofluoromethane	7.914	113	145954	49.62	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.24%
50) Toluene-d8	10.338	98	605822	54.12	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.24%
62) 4-Bromofluorobenzene	12.626	95	250356	64.52	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	129.04%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121021\
Data File : VD071254.D
Acq On : 10 Dec 2021 16:23
Operator : VA/SY
Sample : M4990-04
Misc : 6.80G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
R3-7

Quant Time: Dec 11 04:40:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

