

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121021\
 Data File : VD071259.D
 Acq On : 10 Dec 2021 18:43
 Operator : VA/SY
 Sample : M5018-29
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP5

Quant Time: Dec 11 04:42:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	68232	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	119604	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	138934	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	71751	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	55179	68.92	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	137.84%
35) Dibromofluoromethane	7.914	113	43657	57.32	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.64%
50) Toluene-d8	10.338	98	152906	52.76	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.52%
62) 4-Bromofluorobenzene	12.620	95	69698	69.37	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	138.74%
Target Compounds						
20) Methylene Chloride	4.962	84	6464	4.17	ug/l #	Qvalue 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121021\
Data File : VD071259.D
Acq On : 10 Dec 2021 18:43
Operator : VA/SY
Sample : M5018-29
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP5

Quant Time: Dec 11 04:42:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

