

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121121\
 Data File : VD071273.D
 Acq On : 11 Dec 2021 16:48
 Operator : VA/SY
 Sample : M5018-39
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP6

Quant Time: Dec 12 17:54:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	287060	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	491566	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	535083	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	262321	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	146979	43.636	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	87.280%
35) Dibromofluoromethane	7.914	113	151565	48.420	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.840%
50) Toluene-d8	10.338	98	627856	52.707	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.420%
62) 4-Bromofluorobenzene	12.626	95	259348	62.808	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	125.620%
Target Compounds						
16) Acetone	4.150	43	4236	34.949	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121121\
Data File : VD071273.D
Acq On : 11 Dec 2021 16:48
Operator : VA/SY
Sample : M5018-39
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP6

Quant Time: Dec 12 17:54:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

