

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121121\
 Data File : VD071281.D
 Acq On : 11 Dec 2021 20:34
 Operator : VA/SY
 Sample : M5017-29
 Misc : 5.04G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP3

Quant Time: Dec 12 18:52:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	142130	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.861	114	252384	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	292617	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	153113	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	99933	59.922	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 119.840%	
35) Dibromofluoromethane	7.914	113	86249	53.666	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 107.340%	
50) Toluene-d8	10.332	98	321793	52.615	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 105.220%	
62) 4-Bromofluorobenzene	12.626	95	147430	69.541	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 139.080%	
Target Compounds						Qvalue
16) Acetone	4.156	43	3097	36.928	ug/l	91
20) Methylene Chloride	4.956	84	14328	4.733	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121121\
Data File : VD071281.D
Acq On : 11 Dec 2021 20:34
Operator : VA/SY
Sample : M5017-29
Misc : 5.04G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP3

Quant Time: Dec 12 18:52:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

