

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121223\
 Data File : VD077837.D
 Acq On : 12 Dec 2023 13:03
 Operator : JC/SY
 Sample : 05720-02
 Misc : 2.16G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TT-FRACSEDIMENT-IDWSO-20231207

Quant Time: Dec 12 13:53:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 08 13:33:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	143813	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	284659	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	235047	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	76077	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	70136	41.387	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	82.780%
35) Dibromofluoromethane	7.811	113	85491	42.875	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	85.760%
50) Toluene-d8	10.269	98	352782	44.175	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	88.360%
62) 4-Bromofluorobenzene	12.575	95	77919	32.635	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	65.260%
Target Compounds						
					Qvalue	
44) Trichloroethene	9.034	130	5215	2.211	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121223\
Data File : VD077837.D
Acq On : 12 Dec 2023 13:03
Operator : JC/SY
Sample : 05720-02
Misc : 2.16G/5.0ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TT-FRACSEDIMENT-IDWSO-20231207

Quant Time: Dec 12 13:53:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
Quant Title : SW846 8260
QLast Update : Fri Dec 08 13:33:18 2023
Response via : Initial Calibration

