

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121223\
 Data File : VD077839.D
 Acq On : 12 Dec 2023 13:57
 Operator : JC/SY
 Sample : 05699-01
 Misc : 5.02G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ELM-N7THSTREET-DRUMS-SOIL

Quant Time: Dec 12 14:23:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 08 13:33:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	19436	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.787	114	37519	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.587	117	41589	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	18797	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	15994	69.835	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	139.680%
35) Dibromofluoromethane	7.805	113	13203	50.238	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.480%
50) Toluene-d8	10.269	98	49377	46.910	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	93.820%
62) 4-Bromofluorobenzene	12.575	95	16625	52.829	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	105.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121223\
Data File : VD077839.D
Acq On : 12 Dec 2023 13:57
Operator : JC/SY
Sample : 05699-01
Misc : 5.02G/5.0ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
ELM-N7THSTREET-DRUMS-SOIL

Quant Time: Dec 12 14:23:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
Quant Title : SW846 8260
QLast Update : Fri Dec 08 13:33:18 2023
Response via : Initial Calibration

